

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040424\
 Data File : VY017847.D
 Acq On : 04 Apr 2024 12:24
 Operator : SY/MD
 Sample : P1974-04
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-2

Quant Time: Apr 05 00:45:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	136904	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.709	114	257992	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.508	117	230381	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.440	152	87306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	75703	63.902	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.800%
35) Dibromofluoromethane	7.734	113	83316	55.887	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.780%
50) Toluene-d8	10.197	98	299727	53.823	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.640%
62) 4-Bromofluorobenzene	12.495	95	81564	48.141	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040424\
Data File : VY017847.D
Acq On : 04 Apr 2024 12:24
Operator : SY/MD
Sample : P1974-04
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-2

Quant Time: Apr 05 00:45:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

