

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040424\
 Data File : VY017849.D
 Acq On : 04 Apr 2024 13:11
 Operator : SY/MD
 Sample : P1980-01
 Misc : 5.59g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC1-GRAB-10-11

Quant Time: Apr 05 00:46:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	56330	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.704	114	76107	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.502	117	48522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	11751	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	15457	31.710	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	63.420%
35) Dibromofluoromethane	7.728	113	21326	48.492	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.980%
50) Toluene-d8	10.191	98	82303	50.100	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.200%
62) 4-Bromofluorobenzene	12.496	95	12965	25.940	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	51.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040424\
Data File : VY017849.D
Acq On : 04 Apr 2024 13:11
Operator : SY/MD
Sample : P1980-01
Misc : 5.59g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC1-GRAB-10-11

Quant Time: Apr 05 00:46:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

