

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040424\
 Data File : VY017852.D
 Acq On : 04 Apr 2024 14:21
 Operator : SY/MD
 Sample : P1980-01RE
 Misc : 5.53g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC1-GRAB-10-11RE

Quant Time: Apr 05 00:47:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	124622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	236407	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	212596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	78565	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	71589	66.384	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.760%
35) Dibromofluoromethane	7.728	113	76602	56.075	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.140%
50) Toluene-d8	10.185	98	277397	54.361	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.720%
62) 4-Bromofluorobenzene	12.489	95	70439	45.370	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.740%
Target Compounds						
20) Methylene Chloride	4.722	84	16851	10.224	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040424\
Data File : VY017852.D
Acq On : 04 Apr 2024 14:21
Operator : SY/MD
Sample : P1980-01RE
Misc : 5.53g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC1-GRAB-10-11RE

Quant Time: Apr 05 00:47:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

