

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040521\
 Data File : VY004280.D
 Acq On : 05 Apr 2021 13:34
 Operator : SY/MD
 Sample : M1932-02
 Misc : 6.64g/5mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TS-2

Quant Time: Apr 05 15:21:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	188741	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	311075	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	272116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	118075	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	111909	61.68	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.36%	
35) Dibromofluoromethane	7.728	113	93467	53.92	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.84%	
50) Toluene-d8	10.185	98	375628	57.62	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 115.24%	
62) 4-Bromofluorobenzene	12.483	95	118645	53.10	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.20%	
Target Compounds						
16) Acetone	3.960	43	20814	55.06	ug/l	97
25) 2-Butanone	7.002	43	7275	11.52	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040521\
Data File : VY004280.D
Acq On : 05 Apr 2021 13:34
Operator : SY/MD
Sample : M1932-02
Misc : 6.64g/5mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TS-2

Quant Time: Apr 05 15:21:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

