

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040521\
 Data File : VY004281.D
 Acq On : 05 Apr 2021 13:57
 Operator : SY/MD
 Sample : M1932-03
 Misc : 5.96g/5mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TS-3

Quant Time: Apr 05 15:21:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	194343	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	337137	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	307050	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	129706	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	120082	64.28	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.56%
35) Dibromofluoromethane	7.728	113	103219	54.95	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.90%
50) Toluene-d8	10.185	98	408118	57.76	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.52%
62) 4-Bromofluorobenzene	12.483	95	131613	54.35	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.70%
Target Compounds						
16) Acetone	3.960	43	3707	9.52	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040521\
Data File : VY004281.D
Acq On : 05 Apr 2021 13:57
Operator : SY/MD
Sample : M1932-03
Misc : 5.96g/5mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TS-3

Quant Time: Apr 05 15:21:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

