



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040521\
Data File : VY004274.D
Acq On : 05 Apr 2021 09:20
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y0329215.M
Title : SW846 8260
Last Update : Tue Mar 30 06:38:14 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.1	19971	PASS
75	95	30	60	54.4	51518	PASS
95	95	100	100	100.0	94779	PASS
96	95	5	9	6.6	6280	PASS
173	174	0.00	2	1.0	741	PASS
174	95	50	100	79.2	75069	PASS
175	174	5	9	7.6	5703	PASS
176	174	95	101	98.0	73552	PASS
177	176	5	9	7.0	5155	PASS