

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY040620\
 Data File : VY002238.D
 Acq On : 06 Apr 2020 15:39
 Operator : SY/MD
 Sample : L2169-03
 Misc : 5.03G/5ML/MSVOA Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 96ST-3

Quant Time: Apr 07 05:17:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y032620S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 07:11:05 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	38599	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.70	114	70635	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	60451	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.43	152	25482	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	20958	49.12	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	98.24%	
35) Dibromofluoromethane	7.73	113	18963	47.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.82%	
50) Toluene-d8	10.18	98	85774	52.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.34%	
62) 4-Bromofluorobenzene	12.49	95	25738	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
Target Compounds						
16) Acetone	3.96	43	9699	80.905	ug/l	Qvalue 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY040620\
Data File : VY002238.D
Acq On : 06 Apr 2020 15:39
Operator : SY/MD
Sample : L2169-03
Misc : 5.03G/5ML/MSVOA Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
96ST-3

Quant Time: Apr 07 05:17:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y032620S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 07:11:05 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.80 min Scan# 995

Delta R.T. -0.01 min

Lab File: VY002238.D

Acq: 06 Apr 2020 15:39

Instrument :

MSVOA_Y

ClientSampleId :


96ST-3

Tot Ion:168 Resp: 38599




Ion Ratio Lower Upper

168 100

99 53.1 47.6 71.4



Abundance Ion 167.90 (167.60 to 168.60): VY002238.D (-947) (-)



#16

Acetone

Concen: 80.905 ug/l

RT: 3.96 min Scan# 364

Delta R.T. -0.01 min

Lab File: VY002238.D


Acq: 06 Apr 2020 15:39

Tgt Ion: 43 Resp: 9699



Ion Ratio Lower Upper


43 100

58 29.9 26.3 39.5



Abundance Ion 43.00 (42.70 to 43.70): VY002238.D (-316) (-)









#33
 1,2-Dichloroethane-d4
 Concen: 49.116 ug/l
 RT: 8.15 min Scan# 1052
 Delta R.T. -0.01 min
 Lab File: VY002238.D
 Acq: 06 Apr 2020 15:39




Instrument :
 MSVOA_Y
 ClientSampled :
 96ST-3


Tot Ion: 65 Resp: 20958
 Ion Ratio Lower Upper
 65 100
 67 55.3 0.0 112.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.70 min Scan# 1142
 Delta R.T. -0.01 min
 Lab File: VY002238.D
 Acq: 06 Apr 2020 15:39

Tgt Ion:114 Resp: 70635
 Ion Ratio Lower Upper
 114 100
 63 19.7 0.0 41.4
 88 15.1 0.0 33.0





#35

Dibromofluoromethane

Concen: 47.913 ug/l

RT: 7.73 min Scan# 983

Delta R.T. -0.01 min

Lab File: VY002238.D

Acq: 06 Apr 2020 15:39

Instrument :

MSVOA_Y

ClientSampleId :

96ST-3


Tot Ion:113 Resp: 18963

Ion Ratio Lower Upper

113 100

111 102.7 84.4 126.6




192 18.9 14.6 21.8



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#50

Toluene-d8

Concen: 52.169 ug/l

RT: 10.18 min Scan# 1385

Delta R.T. -0.01 min

Lab File: VY002238.D


Acq: 06 Apr 2020 15:39

Tgt Ion: 98 Resp: 85774

Ion Ratio Lower Upper



98 100


100 65.0 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VY0

Ion 100.00 (99.70 to 100.70): VY









#62
4-Bromofluorobenzene
Concen: 44.760 ug/l
RT: 12.49 min Scan# 1763
Delta R.T. 0.00 min
Lab File: VY002238.D
Acq: 06 Apr 2020 15:39




Instrument :
MSVOA_Y
ClientSampleId :
96ST-3


Tot Ion: 95 Resp: 25738
Ion Ratio Lower Upper
95 100
174 74.8 0.0 143.4
176 71.1 0.0 137.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.49 min Scan# 1600
Delta R.T. -0.01 min
Lab File: VY002238.D
Acq: 06 Apr 2020 15:39

Tgt Ion: 117 Resp: 60451
Ion Ratio Lower Upper
117 100
82 56.8 45.4 68.2
119 31.0 25.3 37.9





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.43 min Scan# 1918

Delta R.T. 0.00 min

Lab File: VY002238.D

Acq: 06 Apr 2020 15:39

Instrument :

MSVOA_Y

ClientSampleId :

96ST-3


Tot Ion:152 Resp: 25482

Ion Ratio Lower Upper

152 100

115 63.6 31.2 93.6

150 161.5 0.0 354.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

