


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv040620\
 Data File : VY002225.D
 Acq On : 06 Apr 2020 09:36
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y032620S.M
 Title : SW846 8260
 Last Update : Fri Mar 27 07:11:05 2020



AutoFind: Scans 1763, 1764, 1765; Background Corrected with Scan 1755

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	29752	PASS
75	95	30	60	51.0	80904	PASS
95	95	100	100	100.0	158613	PASS
96	95	5	9	6.8	10736	PASS
173	174	0.00	2	0.8	861	PASS
174	95	50	100	70.1	111149	PASS
175	174	5	9	7.3	8114	PASS
176	174	95	101	97.2	107984	PASS
177	176	5	9	6.9	7492	PASS