

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040621\
 Data File : VY004295.D
 Acq On : 06 Apr 2021 12:20
 Operator : SY/MD
 Sample : M1934-03
 Misc : 5.80g/5mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-2

Quant Time: Apr 07 00:58:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	182819	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	313238	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	293075	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	125241	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	115587	65.77	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.54%
35) Dibromofluoromethane	7.728	113	97201	55.69	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.38%
50) Toluene-d8	10.185	98	384530	58.58	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	117.16%
62) 4-Bromofluorobenzene	12.483	95	127033	56.46	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040621\
Data File : VY004295.D
Acq On : 06 Apr 2021 12:20
Operator : SY/MD
Sample : M1934-03
Misc : 5.80g/5mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-2

Quant Time: Apr 07 00:58:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

