

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040621\
 Data File : VY004300.D
 Acq On : 06 Apr 2021 15:59
 Operator : SY/MD
 Sample : M1933-13
 Misc : 4.94/5mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BENNETT_BH_05_0-1

Quant Time: Apr 07 01:00:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	179599	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	311518	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	290693	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	128478	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	108671	62.94	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.88%
35) Dibromofluoromethane	7.728	113	94451	54.41	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.82%
50) Toluene-d8	10.185	98	379232	58.09	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	116.18%
62) 4-Bromofluorobenzene	12.483	95	125944	56.29	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.58%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040621\
Data File : VY004300.D
Acq On : 06 Apr 2021 15:59
Operator : SY/MD
Sample : M1933-13
Misc : 4.94/5mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BENNETT_BH_05_0-1

Quant Time: Apr 07 01:00:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

