

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040621\
 Data File : VY004303.D
 Acq On : 06 Apr 2021 17:09
 Operator : SY/MD
 Sample : M1933-05
 Misc : 4.94g/5mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BENNETT_BH_01_2-2.5

Quant Time: Apr 07 01:00:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	173250	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	300273	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	286088	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	129139	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.154	65	106878	64.17	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.34%
35) Dibromofluoromethane	7.728	113	91681	54.80	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.60%
50) Toluene-d8	10.185	98	369273	58.68	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	117.36%
62) 4-Bromofluorobenzene	12.483	95	124190	57.58	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	115.16%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040621\
Data File : VY004303.D
Acq On : 06 Apr 2021 17:09
Operator : SY/MD
Sample : M1933-05
Misc : 4.94g/5mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BENNETT_BH_01_2-2.5

Quant Time: Apr 07 01:00:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

