

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040621\
 Data File : VY004295.D
 Acq On : 06 Apr 2021 12:20
 Operator : SY/MD
 Sample : M1934-03
 Misc : 5.80g/5mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-2

Quant Time: Apr 07 00:58:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	182819	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	313238	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	293075	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	125241	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	115587	65.77	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.54%
35) Dibromofluoromethane	7.728	113	97201	55.69	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.38%
50) Toluene-d8	10.185	98	384530	58.58	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	117.16%
62) 4-Bromofluorobenzene	12.483	95	127033	56.46	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.92%


Target Compounds	Qvalue
-----	-----


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040621\
Data File : VY004295.D
Acq On : 06 Apr 2021 12:20
Operator : SY/MD
Sample : M1934-03
Misc : 5.80g/5mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-2

Quant Time: Apr 07 00:58:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.795 min Scan# 980
Delta R.T. -0.006 min
Lab File: VY004295.D
Acq: 06 Apr 2021 12:20




Instrument :
MSVOA_Y
ClientSampled :
TP-2


Tgt Ion:168 Resp: 182819
Ion Ratio Lower Upper
168 100
99 55.5 42.6 64.0



#33
1,2-Dichloroethane-d4
Concen: 65.77 ug/l
RT: 8.148 min Scan# 1038
Delta R.T. -0.000 min
Lab File: VY004295.D
Acq: 06 Apr 2021 12:20

Tgt Ion: 65 Resp: 115587
Ion Ratio Lower Upper
65 100
67 52.6 0.0 103.2









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. 0.000 min
Lab File: VY004295.D
Acq: 06 Apr 2021 12:20




Instrument :
MSVOA_Y
ClientSampled :
TP-2


Tgt Ion:	114	Resp:	313238
Ion	Ratio	Lower	Upper
114	100		
63	20.4	0.0	40.4
88	15.5	0.0	30.8



#35
Dibromofluoromethane
Concen: 55.69 ug/l
RT: 7.728 min Scan# 969
Delta R.T. -0.000 min
Lab File: VY004295.D
Acq: 06 Apr 2021 12:20

Tgt Ion:	113	Resp:	97201
Ion	Ratio	Lower	Upper
113	100		
111	102.3	82.6	124.0
192	19.7	16.4	24.6









#50
Toluene-d8
Concen: 58.58 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. -0.000 min
Lab File: VY004295.D
Acq: 06 Apr 2021 12:20




Instrument :
MSVOA_Y
ClientSampled :
TP-2


Tgt Ion: 98 Resp: 384530
Ion Ratio Lower Upper
98 100
100 64.8 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 56.46 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY004295.D
Acq: 06 Apr 2021 12:20

Tgt Ion: 95 Resp: 127033
Ion Ratio Lower Upper
95 100
174 84.0 0.0 171.4
176 80.1 0.0 165.6









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.495 min Scan# 1587
Delta R.T. -0.000 min
Lab File: VY004295.D
Acq: 06 Apr 2021 12:20

Instrument :
MSVOA_Y
ClientSampled :
TP-2

Tgt Ion:117 Resp: 293075
Ion Ratio Lower Upper
117 100
82 54.8 42.7 64.1
119 31.8 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY004295.D
Acq: 06 Apr 2021 12:20

Tgt Ion:152 Resp: 125241
Ion Ratio Lower Upper
152 100
115 58.3 29.0 87.1
150 157.4 0.0 347.4

