



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040723\
 Data File : VY013218.D
 Acq On : 07 Apr 2023 16:06
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040723S.M
 Title : SW846 8260
 Last Update : Mon Apr 10 01:29:21 2023



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.5	9900	PASS
75	95	30	60	56.8	23893	PASS
95	95	100	100	100.0	42043	PASS
96	95	5	9	6.8	2844	PASS
173	174	0.00	2	1.5	551	PASS
174	95	50	100	85.5	35933	PASS
175	174	5	9	8.2	2961	PASS
176	174	95	101	95.9	34461	PASS
177	176	5	9	6.9	2376	PASS