

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
 Data File : VY004330.D
 Acq On : 08 Apr 2021 13:12
 Operator : SY/MD
 Sample : M1973-01
 Misc : 5.25G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-02-040721

Quant Time: Apr 09 03:15:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	156608	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	254476	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	218361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	89138	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	88952	59.08	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.16%
35) Dibromofluoromethane	7.728	113	76666	54.07	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.14%
50) Toluene-d8	10.185	98	310796	58.28	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	116.56%
62) 4-Bromofluorobenzene	12.489	95	93970	51.41	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.82%
Target Compounds						
16) Acetone	3.960	43	3006	9.58	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
Data File : VY004330.D
Acq On : 08 Apr 2021 13:12
Operator : SY/MD
Sample : M1973-01
Misc : 5.25G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-02-040721

Quant Time: Apr 09 03:15:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

