

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
 Data File : VY004337.D
 Acq On : 08 Apr 2021 15:55
 Operator : SY/MD
 Sample : M1947-01
 Misc : 5.30G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FS-SB19B(6-7)

Quant Time: Apr 09 03:17:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	175230	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	301146	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	282259	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	129541	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	114047	67.70	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.40%
35) Dibromofluoromethane	7.728	113	92150	54.92	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.84%
50) Toluene-d8	10.185	98	369350	58.52	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	117.04%
62) 4-Bromofluorobenzene	12.483	95	129654	59.94	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.88%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
Data File : VY004337.D
Acq On : 08 Apr 2021 15:55
Operator : SY/MD
Sample : M1947-01
Misc : 5.30G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FS-SB19B(6-7)

Quant Time: Apr 09 03:17:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

