

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
 Data File : VY004340.D
 Acq On : 08 Apr 2021 17:05
 Operator : SY/MD
 Sample : M1947-07
 Misc : 5.76G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FS-SB9(20-21)

Quant Time: Apr 09 03:18:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	173810	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	300573	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	281083	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	128683	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	112444	67.30	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.60%
35) Dibromofluoromethane	7.728	113	92213	55.06	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.12%
50) Toluene-d8	10.185	98	367111	58.28	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	116.56%
62) 4-Bromofluorobenzene	12.483	95	127338	58.98	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	117.96%
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	16226	46.61	ug/l	99
17) Carbon Disulfide	4.210	76	24080	5.82	ug/l	96
95) Naphthalene	15.239	128	113677	20.46	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
Data File : VY004340.D
Acq On : 08 Apr 2021 17:05
Operator : SY/MD
Sample : M1947-07
Misc : 5.76G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FS-SB9(20-21)

Quant Time: Apr 09 03:18:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

