

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
 Data File : VY004345.D
 Acq On : 08 Apr 2021 19:02
 Operator : SY/MD
 Sample : M1979-10
 Misc : 4.65G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-10

Quant Time: Apr 09 03:19:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	181595	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	314314	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	294176	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	134753	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	112423	64.40	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.80%	
35) Dibromofluoromethane	7.728	113	95406	54.48	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.96%	
50) Toluene-d8	10.185	98	385918	58.59	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 117.18%	
62) 4-Bromofluorobenzene	12.483	95	132005	58.47	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 116.94%	
Target Compounds						
16) Acetone	3.960	43	6422	17.66	ug/l	95
20) Methylene Chloride	4.728	84	16622	7.17	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
Data File : VY004345.D
Acq On : 08 Apr 2021 19:02
Operator : SY/MD
Sample : M1979-10
Misc : 4.65G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-10

Quant Time: Apr 09 03:19:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

