

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
 Data File : VY004345.D
 Acq On : 08 Apr 2021 19:02
 Operator : SY/MD
 Sample : M1979-10
 Misc : 4.65G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-10

Quant Time: Apr 09 03:19:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.801	168	181595	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	314314	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	294176	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	134753	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	112423	64.40	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.80%	
35) Dibromofluoromethane	7.728	113	95406	54.48	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.96%	
50) Toluene-d8	10.185	98	385918	58.59	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 117.18%	
62) 4-Bromofluorobenzene	12.483	95	132005	58.47	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 116.94%	
Target Compounds						
16) Acetone	3.960	43	6422	17.66	ug/l	95
20) Methylene Chloride	4.728	84	16622	7.17	ug/l	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
Data File : VY004345.D
Acq On : 08 Apr 2021 19:02
Operator : SY/MD
Sample : M1979-10
Misc : 4.65G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-10

Quant Time: Apr 09 03:19:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.801 min Scan# 981
Delta R.T. 0.000 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02




Instrument :
MSVOA_Y
ClientSampleId :
S-10


Tgt Ion:168 Resp: 181595
Ion Ratio Lower Upper
168 100
99 57.1 42.6 64.0



#16
Acetone
Concen: 17.66 ug/l
RT: 3.960 min Scan# 351
Delta R.T. 0.006 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02

Tgt Ion: 43 Resp: 6422
Ion Ratio Lower Upper
43 100
58 27.9 24.5 36.7









#20
Methylene Chloride
Concen: 7.17 ug/l
RT: 4.728 min Scan# 477
Delta R.T. -0.007 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02

Instrument :
MSVOA_Y
ClientSampleId :
S-10

Tgt Ion: 84 Resp: 16622




Ion	Ratio	Lower	Upper
84	100		
49	136.5	101.5	152.3
51	40.7	31.5	47.3
86	66.6	50.8	76.2




#33
1,2-Dichloroethane-d4
Concen: 64.40 ug/l
RT: 8.155 min Scan# 1039
Delta R.T. 0.006 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02

Tgt Ion: 65 Resp: 112423

Ion	Ratio	Lower	Upper
65	100		
67	52.3	0.0	103.2









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. 0.000 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02




Instrument :
MSVOA_Y
ClientSampled :
S-10


Tgt Ion:	114	Resp:	314314
Ion	Ratio	Lower	Upper
114	100		
63	21.5	0.0	40.4
88	15.9	0.0	30.8



#35
Dibromofluoromethane
Concen: 54.48 ug/l
RT: 7.728 min Scan# 969
Delta R.T. -0.000 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02

Tgt Ion:	113	Resp:	95406
Ion	Ratio	Lower	Upper
113	100		
111	103.2	82.6	124.0
192	19.7	16.4	24.6









#50
Toluene-d8
Concen: 58.59 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. -0.000 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02




Instrument :
MSVOA_Y
ClientSampled :
S-10


Tgt Ion: 98 Resp: 385918
Ion Ratio Lower Upper
98 100
100 65.3 52.5 78.7



#62
4-Bromofluorobenzene
Concen: 58.47 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02

Tgt Ion: 95 Resp: 132005
Ion Ratio Lower Upper
95 100
174 82.3 0.0 171.4
176 80.3 0.0 165.6









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.495 min Scan# 1587
Delta R.T. -0.000 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02

Instrument :
MSVOA_Y
ClientSampled :
S-10

Tgt Ion:117 Resp: 294176

Ion	Ratio	Lower	Upper
117	100		
82	55.7	42.7	64.1
119	32.7	25.7	38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY004345.D
Acq: 08 Apr 2021 19:02

Tgt Ion:152 Resp: 134753

Ion	Ratio	Lower	Upper
152	100		
115	59.6	29.0	87.1
150	159.0	0.0	347.4

