



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040821\
Data File : VY004324.D
Acq On : 08 Apr 2021 09:23
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Title : SW846 8260
Last Update : Tue Mar 30 06:38:14 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.1	17290	PASS
75	95	30	60	56.4	44168	PASS
95	95	100	100	100.0	78283	PASS
96	95	5	9	7.3	5710	PASS
173	174	0.00	2	1.3	833	PASS
174	95	50	100	81.9	64141	PASS
175	174	5	9	7.7	4914	PASS
176	174	95	101	95.8	61429	PASS
177	176	5	9	7.4	4550	PASS