

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY040920\
 Data File : VY002267.D
 Acq On : 09 Apr 2020 11:59
 Operator : SY/MD
 Sample : L2205-01
 Misc : 8.42G/5ML/MSVOA Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JC-02-04082020-A

Quant Time: Apr 10 09:00:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y040820S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 08 15:15:02 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.81	168	108493	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.71	114	205064	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.50	117	169283	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.44	152	62158	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.16	65	57490	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
35) Dibromofluoromethane	7.74	113	58489	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
50) Toluene-d8	10.19	98	239714	46.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.88%	
62) 4-Bromofluorobenzene	12.49	95	72639	43.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.22%	
Target Compounds						
16) Acetone	3.96	43	3140	9.881	ug/l	90
69) o-Xylene	12.04	106	6295	2.779	ug/l	99
80) 1,3,5-Trimethylbenzene	12.82	105	43689	10.818	ug/l	99
84) 1,2,4-Trimethylbenzene	13.13	105	16823	4.172	ug/l	97
85) sec-Butylbenzene	13.26	105	8191	1.625	ug/l	90
86) p-Isopropyltoluene	13.38	119	17759	4.095	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY040920\
Data File : VY002267.D
Acq On : 09 Apr 2020 11:59
Operator : SY/MD
Sample : L2205-01
Misc : 8.42G/5ML/MSVOA Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JC-02-04082020-A

Quant Time: Apr 10 09:00:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y040820S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 08 15:15:02 2020
Response via : Initial Calibration

