


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv040920\  
 Data File : VY002262.D  
 Acq On : 09 Apr 2020 09:31  
 Operator : SY/MD  
 Sample : BFB  
 Misc : 5.00G/5ML/MSVOA Y/SOIL  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_Y\METHODS\82Y040820S.M  
 Title : SW846 8260  
 Last Update : Wed Apr 08 15:15:02 2020



AutoFind: Scans 1763, 1764, 1765; Background Corrected with Scan 1755

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 19.0      | 31810   | PASS             |
| 75          | 95           | 30           | 60           | 50.8      | 85082   | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 167488  | PASS             |
| 96          | 95           | 5            | 9            | 6.7       | 11169   | PASS             |
| 173         | 174          | 0.00         | 2            | 0.8       | 900     | PASS             |
| 174         | 95           | 50           | 100          | 68.0      | 113952  | PASS             |
| 175         | 174          | 5            | 9            | 7.3       | 8274    | PASS             |
| 176         | 174          | 95           | 101          | 95.4      | 108698  | PASS             |
| 177         | 176          | 5            | 9            | 7.0       | 7611    | PASS             |