

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040921\
 Data File : VY004357.D
 Acq On : 09 Apr 2021 14:59
 Operator : SY/MD
 Sample : M1979-09RE
 Misc : 4.75G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-9RE

Quant Time: Apr 10 02:02:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	167456	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	289755	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	288552	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	141672	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	101363	62.97	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.94%	
35) Dibromofluoromethane	7.728	113	87387	54.13	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.26%	
50) Toluene-d8	10.185	98	365367	60.17	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 120.34%	
62) 4-Bromofluorobenzene	12.483	95	162374	78.02	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 156.04%#	
Target Compounds						
16) Acetone	3.960	43	5142	15.33	ug/l	Qvalue 93
85) sec-Butylbenzene	13.258	105	21537	2.14	ug/l	# 48
89) n-Butylbenzene	13.697	91	38983	4.54	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040921\
Data File : VY004357.D
Acq On : 09 Apr 2021 14:59
Operator : SY/MD
Sample : M1979-09RE
Misc : 4.75G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-9RE

Quant Time: Apr 10 02:02:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

