

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040921\
 Data File : VY004362.D
 Acq On : 09 Apr 2021 16:55
 Operator : SY/MD
 Sample : M1967-05
 Misc : 3.52G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SED-1C

Quant Time: Apr 10 02:04:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	251033	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	435153	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	396240	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	173874	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	149644	62.01	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.02%
35) Dibromofluoromethane	7.728	113	130487	53.82	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.64%
50) Toluene-d8	10.185	98	529827	58.10	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	116.20%
62) 4-Bromofluorobenzene	12.483	95	179034	57.28	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	114.56%
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	83273	165.63	ug/l	99
52) Toluene	10.252	92	110515	15.27	ug/l	94
84) 1,2,4-Trimethylbenzene	13.123	105	19505	1.88	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040921\
Data File : VY004362.D
Acq On : 09 Apr 2021 16:55
Operator : SY/MD
Sample : M1967-05
Misc : 3.52G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SED-1C

Quant Time: Apr 10 02:04:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

