

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040921\
 Data File : VY004369.D
 Acq On : 09 Apr 2021 19:38
 Operator : SY/MD
 Sample : M1993-02
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T-1

Quant Time: Apr 10 02:06:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	219924	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	376835	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	349843	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	160689	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	133540	63.16	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.32%
35) Dibromofluoromethane	7.728	113	56871	27.09	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	54.18%
50) Toluene-d8	10.185	98	462846	58.61	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	117.22%
62) 4-Bromofluorobenzene	12.483	95	159388	58.89	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	117.78%
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	23399	53.12	ug/l	98
52) Toluene	10.246	92	11788	1.88	ug/l	97
84) 1,2,4-Trimethylbenzene	13.123	105	18533	1.93	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040921\
Data File : VY004369.D
Acq On : 09 Apr 2021 19:38
Operator : SY/MD
Sample : M1993-02
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T-1

Quant Time: Apr 10 02:06:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

