

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040921\
 Data File : VY004351.D
 Acq On : 09 Apr 2021 11:52
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 10 02:00:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.433	0.388	10.4	77	0.00
3 P	Chloromethane	0.584	0.572	2.1	84	0.00
4 C	Vinyl Chloride	0.651	0.687	-5.5#	92	0.00
5 T	Bromomethane	0.358	0.299	16.5	74	0.00
6 T	Chloroethane	0.422	0.386	8.5	80	0.00
7 T	Trichlorofluoromethane	0.847	0.809	4.5	83	0.00
8 T	Diethyl Ether	0.272	0.278	-2.2	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.461	0.454	1.5	86	0.00
10 T	Methyl Iodide	0.533	0.508	4.7	82	0.00
11 T	Tert butyl alcohol	0.039	0.055	-41.0#	115	-0.01
12 CM	1,1-Dichloroethene	0.452	0.436	3.5#	83	0.00
13 T	Acrolein	0.037	0.028	24.3	80	0.00
14 T	Allyl chloride	0.721	0.847	-17.5	101	0.00
15 T	Acrylonitrile	0.129	0.161	-24.8	103	0.00
16 T	Acetone	0.100	0.105	-5.0	92	0.00
17 T	Carbon Disulfide	1.189	1.236	-4.0	86	0.00
18 T	Methyl Acetate	0.321	0.418	-30.2#	109	0.00
19 T	Methyl tert-butyl Ether	1.336	1.395	-4.4	89	0.00
20 T	Methylene Chloride	0.554	0.523	5.6	91	0.00
21 T	trans-1,2-Dichloroethene	0.508	0.494	2.8	82	0.00
22 T	Diisopropyl ether	1.583	1.864	-17.8	102	0.00
23 T	Vinyl Acetate	1.065	1.257	-18.0	99	0.00
24 P	1,1-Dichloroethane	0.862	0.931	-8.0	93	0.00
25 T	2-Butanone	0.167	0.205	-22.8	100	0.00
26 T	2,2-Dichloropropane	0.807	0.858	-6.3	92	0.00
27 T	cis-1,2-Dichloroethene	0.575	0.564	1.9	84	0.00
28 T	Bromochloromethane	0.351	0.410	-16.8	108	0.00
29 T	Tetrahydrofuran	0.118	0.151	-28.0#	107	0.00
30 C	Chloroform	0.917	0.926	-1.0#	87	0.00
31 T	Cyclohexane	0.849	0.885	-4.2	95	0.00
32 T	1,1,1-Trichloroethane	0.839	0.827	1.4	85	0.00
33 S	1,2-Dichloroethane-d4	0.481	0.489	-1.7	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.279	0.256	8.2	83	0.00
36 T	1,1-Dichloropropene	0.456	0.457	-0.2	89	0.00
37 T	Ethyl Acetate	0.268	0.310	-15.7	102	0.00
38 T	Carbon Tetrachloride	0.397	0.425	-7.1	91	0.00
39 T	Methylcyclohexane	0.565	0.573	-1.4	89	0.00
40 TM	Benzene	1.294	1.296	-0.2	89	0.00
41 T	Methacrylonitrile	0.148	0.143	3.4	88	-0.01
42 TM	1,2-Dichloroethane	0.407	0.406	0.2	87	0.00
43 T	Isopropyl Acetate	0.507	0.603	-18.9	102	0.00
44 TM	Trichloroethene	0.412	0.356	13.6	78	0.00
45 C	1,2-Dichloropropane	0.320	0.341	-6.6#	94	0.00
46 T	Dibromomethane	0.188	0.185	1.6	85	0.00
47 T	Bromodichloromethane	0.449	0.456	-1.6	88	0.00
48 T	Methyl methacrylate	0.230	0.282	-22.6	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040921\
 Data File : VY004351.D
 Acq On : 09 Apr 2021 11:52
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 10 02:00:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	106	0.00
50 S	Toluene-d8	1.048	0.999	4.7	88	0.00
51 T	4-Methyl-2-Pentanone	0.258	0.322	-24.8	106	0.00
52 CM	Toluene	0.832	0.823	1.1#	87	0.00
53 T	t-1,3-Dichloropropene	0.478	0.503	-5.2	91	0.00
54 T	cis-1,3-Dichloropropene	0.534	0.562	-5.2	91	0.00
55 T	1,1,2-Trichloroethane	0.258	0.261	-1.2	89	0.00
56 T	Ethyl methacrylate	0.353	0.387	-9.6	93	0.00
57 T	1,3-Dichloropropane	0.444	0.467	-5.2	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.151	0.102	32.5#	66	0.00
59 T	2-Hexanone	0.178	0.216	-21.3	103	0.00
60 T	Dibromochloromethane	0.309	0.308	0.3	86	0.00
61 T	1,2-Dibromoethane	0.254	0.253	0.4	87	0.00
62 S	4-Bromofluorobenzene	0.359	0.345	3.9	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.492	0.408	17.1	75	0.00
65 PM	Chlorobenzene	0.997	0.970	2.7	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.350	2.0	86	0.00
67 C	Ethyl Benzene	1.811	1.820	-0.5#	88	0.00
68 T	m/p-Xylenes	0.695	0.680	2.2	86	0.00
69 T	o-Xylene	0.651	0.625	4.0	85	0.00
70 T	Styrene	1.082	1.068	1.3	86	0.00
71 P	Bromoform	0.203	0.211	-3.9	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.653	3.632	0.6	88	0.00
74 T	N-amyl acetate	1.068	1.273	-19.2	102	0.00
75 P	1,1,2,2-Tetrachloroethane	0.676	0.714	-5.6	91	0.00
76 T	1,2,3-Trichloropropane	0.454	0.519	-14.3	98	0.00
77 T	Bromobenzene	0.855	0.803	6.1	83	0.00
78 T	n-propylbenzene	4.226	4.355	-3.1	91	0.00
79 T	2-Chlorotoluene	2.322	2.354	-1.4	89	0.00
80 T	1,3,5-Trimethylbenzene	2.980	2.975	0.2	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.246	0.285	-15.9	101	0.00
82 T	4-Chlorotoluene	2.444	2.472	-1.1	90	0.00
83 T	tert-Butylbenzene	2.575	2.580	-0.2	90	0.00
84 T	1,2,4-Trimethylbenzene	2.989	2.981	0.3	88	0.00
85 T	sec-Butylbenzene	3.545	3.617	-2.0	90	0.00
86 T	p-Isopropyltoluene	3.298	3.284	0.4	88	0.00
87 T	1,3-Dichlorobenzene	1.596	1.521	4.7	84	0.00
88 T	1,4-Dichlorobenzene	1.576	1.519	3.6	85	0.00
89 T	n-Butylbenzene	3.032	3.134	-3.4	90	0.00
90 T	Hexachloroethane	0.229	0.445	-94.3#	157#	0.00
91 T	1,2-Dichlorobenzene	1.432	1.346	6.0	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.126	0.122	3.2	89	0.00
93 T	1,2,4-Trichlorobenzene	1.018	0.899	11.7	78	0.00
94 T	Hexachlorobutadiene	0.541	0.484	10.5	78	0.00
95 T	Naphthalene	2.159	2.016	6.6	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040921\
Data File : VY004351.D
Acq On : 09 Apr 2021 11:52
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 10 02:00:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.930	0.791	14.9	77	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6