

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
 Data File : VY017905.D
 Acq On : 11 Apr 2024 12:50
 Operator : SY/MD
 Sample : P2072-03
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-041024

Quant Time: Apr 12 04:59:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	9405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	13767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	11444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	5114	62.837	ug/l	0.00
Spiked Amount 50.000	Range 50	- 163	Recovery	=	125.680%	
35) Dibromofluoromethane	7.722	113	4561	57.334	ug/l	0.00
Spiked Amount 50.000	Range 54	- 147	Recovery	=	114.660%	
50) Toluene-d8	10.185	98	15159	51.013	ug/l	0.00
Spiked Amount 50.000	Range 58	- 134	Recovery	=	102.020%	
62) 4-Bromofluorobenzene	12.489	95	4221	46.687	ug/l	0.00
Spiked Amount 50.000	Range 30	- 143	Recovery	=	93.380%	
Target Compounds						Qvalue
3) Chloromethane	2.107	50	321	2.615	ug/l #	43
5) Bromomethane	2.662	94	1087	8.380	ug/l	82
10) Methyl Iodide	4.107	142	417	4.087	ug/l #	58
16) Acetone	3.936	43	183	10.878	ug/l #	36
31) Cyclohexane	7.795	56	195	1.360	ug/l #	52
36) 1,1-Dichloropropene	7.789	75	744	6.410	ug/l #	40
38) Carbon Tetrachloride	7.789	117	819	6.797	ug/l #	13
51) 4-Methyl-2-Pentanone	10.069	43	49	Below Cal	#	31
55) 1,1,2-Trichloroethane	10.624	97	74	1.176	ug/l #	16
59) 2-Hexanone	10.807	43	49	1.070	ug/l #	19
68) m/p-Xylenes	11.721	106	200	1.307	ug/l #	49
69) o-Xylene	12.032	106	1353	10.491	ug/l	95
70) Styrene	12.044	104	91	Below Cal	#	1
73) Isopropylbenzene	12.337	105	488	1.442	ug/l	94
74) N-amyl acetate	12.123	43	1010	17.669	ug/l #	73
75) 1,1,2,2-Tetrachloroethane	12.544	83	539	9.209	ug/l #	59
76) 1,2,3-Trichloropropane	12.483	75	2351	53.437	ug/l #	100
78) n-propylbenzene	12.678	91	1210	2.931	ug/l	83
79) 2-Chlorotoluene	12.776	91	844	3.732	ug/l #	39
80) 1,3,5-Trimethylbenzene	12.818	105	5577	21.121	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.483	75	2351	104.853	ug/l #	19
82) 4-Chlorotoluene	12.867	91	360	1.537	ug/l #	39
84) 1,2,4-Trimethylbenzene	13.129	105	1625	6.179	ug/l	93
85) sec-Butylbenzene	13.257	105	1825	5.065	ug/l	93
86) p-Isopropyltoluene	13.373	119	2025	6.724	ug/l	89
88) 1,4-Dichlorobenzene	13.452	146	217	1.418	ug/l #	1
89) n-Butylbenzene	13.696	91	1482	5.348	ug/l #	64
90) Hexachloroethane	13.977	117	1026	17.132	ug/l #	12
91) 1,2-Dichlorobenzene	13.745	146	158	1.192	ug/l #	1
92) 1,2-Dibromo-3-Chloropr...	14.343	75	30	3.793	ug/l #	1
93) 1,2,4-Trichlorobenzene	15.013	180	168	2.314	ug/l #	1
95) Naphthalene	15.245	128	1053	11.042	ug/l #	91
96) 1,2,3-Trichlorobenzene	15.428	180	350	5.708	ug/l #	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
Data File : VY017905.D
Acq On : 11 Apr 2024 12:50
Operator : SY/MD
Sample : P2072-03
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024

Quant Time: Apr 12 04:59:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50




Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024


Tgt Ion:168 Resp: 9405
Ion Ratio Lower Upper
168 100
99 54.6 39.1 58.7



#3
Chloromethane
Concen: 2.615 ug/l
RT: 2.107 min Scan# 47
Delta R.T. -0.006 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50

Tgt Ion: 50 Resp: 321
Ion Ratio Lower Upper
50 100
52 0.0 25.6 38.4#









#5
Bromomethane
Concen: 8.380 ug/l
RT: 2.662 min Scan# 11
Delta R.T. 0.012 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50




Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024


Tgt Ion: 94 Resp: 1087
Ion Ratio Lower Upper
94 100
96 109.5 74.1 111.1



#10
Methyl Iodide
Concen: 4.087 ug/l
RT: 4.107 min Scan# 375
Delta R.T. 0.019 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50

Tgt Ion: 142 Resp: 417
Ion Ratio Lower Upper
142 100
127 13.2 32.8 49.2#
141 0.0 11.4 17.2#





#16

Acetone

Concen: 10.878 ug/l

RT: 3.936 min Scan# 34

Delta R.T. -0.006 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50

Instrument :

MSVOA_Y

ClientSampleId :





HR-02-041024

Tgt Ion: 43 Resp: 183

Ion Ratio Lower Upper

43 100

58 0.0 31.7 47.5#



#31

Cyclohexane

Concen: 1.360 ug/l

RT: 7.795 min Scan# 980

Delta R.T. 0.006 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50




Tgt Ion: 56 Resp: 195


Ion Ratio Lower Upper

56 100

69 122.6 28.7 43.1#

84 86.8 77.7 116.5









#33
 1,2-Dichloroethane-d4
 Concen: 62.837 ug/l
 RT: 8.142 min Scan# 1037
 Delta R.T. -0.000 min
 Lab File: VY017905.D
 Acq: 11 Apr 2024 12:50




Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-041024


Tgt Ion: 65 Resp: 5114
 Ion Ratio Lower Upper
 65 100
 67 48.5 0.0 109.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.691 min Scan# 1127
 Delta R.T. -0.000 min
 Lab File: VY017905.D
 Acq: 11 Apr 2024 12:50

Tgt Ion: 114 Resp: 13767
 Ion Ratio Lower Upper
 114 100
 63 19.2 0.0 35.0
 88 13.5 0.0 27.2









#35
Dibromofluoromethane
Concen: 57.334 ug/l
RT: 7.722 min Scan# 911
Delta R.T. 0.006 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50




Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024


Tgt Ion:113 Resp: 4561
Ion Ratio Lower Upper
113 100
111 106.3 82.2 123.4
192 21.2 15.9 23.9



#36
1,1-Dichloropropene
Concen: 6.410 ug/l
RT: 7.789 min Scan# 979
Delta R.T. -0.128 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50


Tgt Ion: 75 Resp: 744
Ion Ratio Lower Upper
75 100
110 0.0 19.0 57.0#
77 0.0 25.3 37.9#








#38
Carbon Tetrachloride
Concen: 6.797 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.116 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024







Tgt Ion:117 Resp: 819
Ion Ratio Lower Upper
117 100
119 0.0 75.8 113.6#
121 0.0 24.0 36.0#



#50
Toluene-d8
Concen: 51.013 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. -0.000 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50

Tgt Ion: 98 Resp: 15159
Ion Ratio Lower Upper
98 100
100 61.0 52.0 78.0





#55

1,1,2-Trichloroethane

Concen: 1.176 ug/l

RT: 10.624 min Scan# 1444

Delta R.T. -0.024 min

Lab File: VY017905.D


Acq: 11 Apr 2024 12:50

Instrument :

MSVOA_Y

ClientSampleId :

HR-02-041024



Tgt Ion: 97 Resp: 74




Ion Ratio Lower Upper

97 100

83 0.0 67.5 101.3#

85 0.0 42.9 64.3#

99 0.0 46.6 70.0#



#59

2-Hexanone

Concen: 1.070 ug/l

RT: 10.807 min Scan# 1474

Delta R.T. -0.030 min

Lab File: VY017905.D




Acq: 11 Apr 2024 12:50


Tgt Ion: 43 Resp: 49

Ion Ratio Lower Upper

43 100

58 0.0 31.4 94.0#





#62

4-Bromofluorobenzene

Concen: 46.687 ug/l

RT: 12.489 min Scan# 1750

Delta R.T. -0.000 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50

Instrument :

MSVOA_Y

ClientSampleId :

HR-02-041024





Tgt Ion: 95 Resp: 4221

Ion Ratio Lower Upper

95 100

174 96.0 0.0 175.6

176 91.2 0.0 171.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.495 min Scan# 1587

Delta R.T. -0.000 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50




Tgt Ion: 117 Resp: 11444


Ion Ratio Lower Upper

117 100

82 50.3 42.4 63.6

119 32.9 25.9 38.9









#68
m/p-Xylenes
Concen: 1.307 ug/l
RT: 11.721 min Scan# 1624
Delta R.T. 0.012 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50




Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024


Tgt Ion:106 Resp: 200
Ion Ratio Lower Upper
106 100
91 270.5 156.0 234.0#



#69
o-Xylene
Concen: 10.491 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50

Tgt Ion:106 Resp: 1353
Ion Ratio Lower Upper
106 100
91 200.1 103.6 310.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.428 min Scan# 1904

Delta R.T. -0.006 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50

Instrument :

MSVOA_Y

ClientSampleId :

HR-02-041024





Tgt Ion:152 Resp: 4828

Ion Ratio Lower Upper

152 100

115 66.0 28.2 84.7

150 150.6 0.0 345.6



#73

Isopropylbenzene

Concen: 1.442 ug/l

RT: 12.337 min Scan# 1725

Delta R.T. -0.000 min

Lab File: VY017905.D




Acq: 11 Apr 2024 12:50


Tgt Ion:105 Resp: 488

Ion Ratio Lower Upper

105 100

120 23.4 13.4 40.1





#74

N-amyl acetate

Concen: 17.669 ug/l

RT: 12.123 min Scan# 1

Delta R.T. -0.025 min

Lab File: VY017905.D


Acq: 11 Apr 2024 12:50

Instrument :

MSVOA_Y

ClientSampleId :

HR-02-041024



Tgt Ion: 43 Resp: 1010




Ion Ratio Lower Upper

43 100

70 45.0 41.4 62.2

55 47.2 23.3 34.9#

61 0.0 22.6 33.8#



#75

1,1,2,2-Tetrachloroethane

Concen: 9.209 ug/l

RT: 12.544 min Scan# 1759

Delta R.T. -0.043 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50




Tgt Ion: 83 Resp: 539


Ion Ratio Lower Upper

83 100

131 0.0 5.0 14.9#

85 30.2 32.0 96.0#









#76
1,2,3-Trichloropropane
Concen: 53.437 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.152 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50




Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024


Tgt Ion: 75 Resp: 2351
Ion Ratio Lower Upper
75 100
77 5.2 0.0 0.0#



#78
n-propylbenzene
Concen: 2.931 ug/l
RT: 12.678 min Scan# 1781
Delta R.T. -0.000 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50

Tgt Ion: 91 Resp: 1210
Ion Ratio Lower Upper
91 100
120 14.5 11.5 34.4









#79
2-Chlorotoluene
Concen: 3.732 ug/l
RT: 12.776 min Scan# 1
Delta R.T. 0.012 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50




Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024


Tgt Ion: 91 Resp: 844
Ion Ratio Lower Upper
91 100
126 0.0 17.8 53.5#



#80
1,3,5-Trimethylbenzene
Concen: 21.121 ug/l
RT: 12.818 min Scan# 1804
Delta R.T. -0.000 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50

Tgt Ion: 105 Resp: 5577
Ion Ratio Lower Upper
105 100
120 49.9 25.1 75.2









#81
trans-1,4-Dichloro-2-butene
Concen: 104.853 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.097 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50




Instrument :
MSVOA_Y
ClientSampleId :
HR-02-041024


Tgt Ion: 75 Resp: 2351
Ion Ratio Lower Upper
75 100
53 4.8 66.8 100.2#
89 0.0 35.8 53.6#



#82
4-Chlorotoluene
Concen: 1.537 ug/l
RT: 12.867 min Scan# 1812
Delta R.T. 0.006 min
Lab File: VY017905.D
Acq: 11 Apr 2024 12:50

Tgt Ion: 91 Resp: 360
Ion Ratio Lower Upper
91 100
126 0.0 17.6 52.9#





#84

1,2,4-Trimethylbenzene

Concen: 6.179 ug/l

RT: 13.129 min Scan# 1854

Delta R.T. 0.006 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50

Instrument :

MSVOA_Y

ClientSampleId :





HR-02-041024

Tgt Ion:105 Resp: 1625

Ion Ratio Lower Upper

105 100

120 41.6 23.2 69.6



#85

sec-Butylbenzene

Concen: 5.065 ug/l

RT: 13.257 min Scan# 1876

Delta R.T. -0.000 min

Lab File: VY017905.D




Acq: 11 Apr 2024 12:50


Tgt Ion:105 Resp: 1825

Ion Ratio Lower Upper

105 100

134 23.1 10.1 30.1





#86

p-Isopropyltoluene

Concen: 6.724 ug/l

RT: 13.373 min Scan# 119

Delta R.T. -0.000 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50

Instrument :

MSVOA_Y

ClientSampleId :

HR-02-041024





Tgt Ion:119 Resp: 2025

Ion Ratio Lower Upper

119 100

134 30.5 13.1 39.3

91 28.0 10.7 32.1



#88

1,4-Dichlorobenzene

Concen: 1.418 ug/l

RT: 13.452 min Scan# 1908

Delta R.T. -0.000 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50




Tgt Ion:146 Resp: 217


Ion Ratio Lower Upper

146 100

111 319.4 18.4 55.4#

148 23.0 32.4 97.2#





#89

n-Butylbenzene

Concen: 5.348 ug/l

RT: 13.696 min Scan# 1948

Delta R.T. -0.006 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50

Instrument :

MSVOA_Y

ClientSampleId :

HR-02-041024





Tgt Ion: 91 Resp: 1482

Ion Ratio Lower Upper

91 100

92 73.4 26.6 79.7

134 0.0 12.9 38.7#



#90

Hexachloroethane

Concen: 17.132 ug/l

RT: 13.977 min Scan# 1994

Delta R.T. 0.012 min

Lab File: VY017905.D




Acq: 11 Apr 2024 12:50


Tgt Ion: 117 Resp: 1026

Ion Ratio Lower Upper

117 100

201 0.0 37.4 112.2#









#91
 1,2-Dichlorobenzene
 Concen: 1.192 ug/l
 RT: 13.745 min Scan# 1956
 Delta R.T. -0.000 min
 Lab File: VY017905.D
 Acq: 11 Apr 2024 12:50




Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-041024


Tgt Ion:	146	Resp:	158
Ion	Ratio	Lower	Upper
146	100		
111	363.9	19.9	59.7#
148	40.5	31.8	95.4



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 3.793 ug/l
 RT: 14.343 min Scan# 2054
 Delta R.T. -0.018 min
 Lab File: VY017905.D
 Acq: 11 Apr 2024 12:50

Tgt Ion:	75	Resp:	30
Ion	Ratio	Lower	Upper
75	100		
155	0.0	48.6	145.8#
157	0.0	62.7	188.2#





#93

1,2,4-Trichlorobenzene

Concen: 2.314 ug/l

RT: 15.013 min Scan# 2164

Delta R.T. -0.000 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50

Instrument :

MSVOA_Y

ClientSampleId :

HR-02-041024





Tgt Ion:180 Resp: 168

Ion Ratio Lower Upper

180 100

182 139.3 47.9 143.8

145 579.2 14.2 42.6#



#95

Naphthalene

Concen: 11.042 ug/l

RT: 15.245 min Scan# 2202

Delta R.T. -0.000 min

Lab File: VY017905.D

Acq: 11 Apr 2024 12:50




Tgt Ion:128 Resp: 1053


Ion Ratio Lower Upper

128 100

127 18.3 10.6 16.0#


129 12.3 8.6 13.0





#96
 1,2,3-Trichlorobenzene
 Concen: 5.708 ug/l
 RT: 15.428 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VY017905.D
 Acq: 11 Apr 2024 12:50

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-041024



Tgt Ion:180 Resp: 350
 Ion Ratio Lower Upper
 180 100
 182 40.9 47.6 142.9#
 145 268.3 15.4 46.1#

