

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
 Data File : VY017906.D
 Acq On : 11 Apr 2024 13:14
 Operator : SY/MD
 Sample : P2072-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-01-041024

Quant Time: Apr 12 04:59:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.79
34) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/l	-8.69
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-11.50
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.43
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0	0.000	ug/l	
Spiked Amount	50.000	Range	58 - 134	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	30 - 143	Recovery	=	0.000%#


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
Data File : VY017906.D
Acq On : 11 Apr 2024 13:14
Operator : SY/MD
Sample : P2072-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-041024

Quant Time: Apr 12 04:59:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration







#1
Pentafluorobenzene
Concen: 0.000 ug/l
Expected RT: 7.79 min

Lab File: VY017906.D
Acq: 11 Apr 2024 13:14

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-041024


Tgt Ion: 168
Sig Exp Ratio
168 100
99 48.9




#33
1,2-Dichloroethane-d4
Concen: 0.000 ug/l
Expected RT: 8.14 min

Lab File: VY017906.D
Acq: 11 Apr 2024 13:14

Tgt Ion: 65
Sig Exp Ratio
65 100
67 54.8







#34
 1,4-Difluorobenzene
 Concen: 0.000 ug/l
 Expected RT: 8.69 min
 Lab File: VY017906.D
 Acq: 11 Apr 2024 13:14

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-01-041024

Tgt Ion: 114


Sig	Exp Ratio
114	100
63	17.5
88	13.6




#35
 Dibromofluoromethane
 Concen: 0.000 ug/l
 Expected RT: 7.72 min
 Lab File: VY017906.D
 Acq: 11 Apr 2024 13:14

Tgt Ion: 113

Sig	Exp Ratio
113	100
111	102.8
192	19.9







#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 10.18 min

Lab File: VY017906.D
Acq: 11 Apr 2024 13:14

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-041024


Tgt Ion: 98
Sig Exp Ratio
98 100
100 65.0




#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 12.49 min

Lab File: VY017906.D
Acq: 11 Apr 2024 13:14

Tgt Ion: 95
Sig Exp Ratio
95 100
174 87.8
176 85.7







#63
Chlorobenzene-d5
Concen: 0.000 ug/l
Expected RT: 11.50 min

Lab File: VY017906.D
Acq: 11 Apr 2024 13:14

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-041024

Tgt Ion: 117
Sig Exp Ratio
117 100
82 53.0
119 32.4



#72
1,4-Dichlorobenzene-d4
Concen: 0.000 ug/l
Expected RT: 13.43 min

Lab File: VY017906.D
Acq: 11 Apr 2024 13:14

Tgt Ion: 152
Sig Exp Ratio
152 100
115 56.5
150 172.8

