

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
 Data File : VY017901.D
 Acq On : 11 Apr 2024 10:55
 Operator : SY/MD
 Sample : VY0411SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0411SBL01

Quant Time: Apr 12 04:56:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	119363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	221699	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	199539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	72356	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	68414	66.235	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.480%
35) Dibromofluoromethane	7.722	113	71590	55.883	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.760%
50) Toluene-d8	10.185	98	261018	54.545	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.080%
62) 4-Bromofluorobenzene	12.489	95	66402	45.607	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
Data File : VY017901.D
Acq On : 11 Apr 2024 10:55
Operator : SY/MD
Sample : VY0411SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0411SBL01

Quant Time: Apr 12 04:56:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

