

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
 Data File : VY017910.D
 Acq On : 11 Apr 2024 14:50
 Operator : SY/MD
 Sample : P2070-09
 Misc : 3.14g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-5

Quant Time: Apr 12 05:01:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	104451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	194762	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	179055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	67503	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	62393	69.030	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.060%
35) Dibromofluoromethane	7.722	113	65435	58.143	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.280%
50) Toluene-d8	10.185	98	226465	53.870	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.740%
62) 4-Bromofluorobenzene	12.483	95	59705	46.679	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
Data File : VY017910.D
Acq On : 11 Apr 2024 14:50
Operator : SY/MD
Sample : P2070-09
Misc : 3.14g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-5

Quant Time: Apr 12 05:01:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

