

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
 Data File : VY017916.D
 Acq On : 11 Apr 2024 17:12
 Operator : SY/MD
 Sample : P2088-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-04-041124

Quant Time: Apr 12 05:03:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	94506	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	169848	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	149769	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	57294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	54120	66.178	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.360%
35) Dibromofluoromethane	7.722	113	56832	57.906	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.820%
50) Toluene-d8	10.179	98	199680	54.466	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.940%
62) 4-Bromofluorobenzene	12.483	95	51155	45.861	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.720%
Target Compounds						
16) Acetone	3.948	43	2729	16.144	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041124\
Data File : VY017916.D
Acq On : 11 Apr 2024 17:12
Operator : SY/MD
Sample : P2088-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-04-041124

Quant Time: Apr 12 05:03:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

