

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041221\
 Data File : VY004377.D
 Acq On : 12 Apr 2021 11:40
 Operator : SY/MD
 Sample : VSTDIC010
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010

Quant Time: Apr 13 02:50:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 02:48:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	211650	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	337851	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	297999	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	138219	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	22461	9.50	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	19.00%#
35) Dibromofluoromethane	7.728	113	18776	9.61	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	19.22%#
50) Toluene-d8	10.185	98	76051	9.63	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	19.26%#
62) 4-Bromofluorobenzene	12.483	95	25626	9.64	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	19.28%#

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.912	85	19976	9.86	ug/l	94
3) Chloromethane	2.119	50	29547	9.79	ug/l	100
4) Vinyl Chloride	2.259	62	31900	9.79	ug/l	95
5) Bromomethane	2.656	94	15433	10.00	ug/l	97
6) Chloroethane	2.802	64	18234	9.80	ug/l	98
7) Trichlorofluoromethane	3.137	101	39272	10.08	ug/l	99
8) Diethyl Ether	3.546	74	13344	10.39	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.911	101	21212	9.93	ug/l	99
10) Methyl Iodide	4.100	142	25388	9.98	ug/l	99
11) Tert butyl alcohol	4.972	59	13311	53.25	ug/l	100
12) 1,1-Dichloroethene	3.881	96	21410	10.11	ug/l	93
13) Acrolein	3.741	56	15697	57.84	ug/l	99
14) Allyl chloride	4.497	41	40745	10.18	ug/l	98
15) Acrylonitrile	5.186	53	36836	53.46	ug/l	97
16) Acetone	3.960	43	32702	53.29	ug/l	95
17) Carbon Disulfide	4.204	76	67514	10.17	ug/l	98
18) Methyl Acetate	4.491	43	17728	10.51	ug/l	100
19) Methyl tert-butyl Ether	5.234	73	64543	10.34	ug/l	96
20) Methylene Chloride	4.728	84	23903	10.15	ug/l	91
21) trans-1,2-Dichloroethene	5.241	96	23786	9.99	ug/l	96
22) Diisopropyl ether	6.131	45	84551	10.16	ug/l	97
23) Vinyl Acetate	6.076	43	295768	51.93	ug/l	100
24) 1,1-Dichloroethane	6.033	63	43679	10.13	ug/l	99
25) 2-Butanone	7.002	43	50226	54.59	ug/l	99
26) 2,2-Dichloropropane	6.996	77	39356	10.02	ug/l	99
27) cis-1,2-Dichloroethene	7.002	96	27044	10.22	ug/l	99
28) Bromochloromethane	7.350	49	17751	9.49	ug/l	100
29) Tetrahydrofuran	7.356	42	34960	53.86	ug/l	100
30) Chloroform	7.521	83	43846	10.24	ug/l	100
31) Cyclohexane	7.795	56	45684	10.05	ug/l #	86
32) 1,1,1-Trichloroethane	7.710	97	39229	10.09	ug/l	100
36) 1,1-Dichloropropene	7.929	75	35196	10.14	ug/l	98
37) Ethyl Acetate	7.088	43	24111	10.78	ug/l	98
38) Carbon Tetrachloride	7.911	117	28449	9.97	ug/l	93
39) Methylcyclohexane	9.191	83	43106	10.12	ug/l	93
40) Benzene	8.167	78	98654	10.13	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041221\
 Data File : VY004377.D
 Acq On : 12 Apr 2021 11:40
 Operator : SY/MD
 Sample : VSTDIC010
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010

Quant Time: Apr 13 02:50:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 02:48:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.362	41	32625	21.87	ug/l #	100
42) 1,2-Dichloroethane	8.246	62	32021	10.46	ug/l	98
43) Isopropyl Acetate	8.283	43	44299	10.61	ug/l	99
44) Trichloroethene	8.947	130	27271	10.20	ug/l	100
45) 1,2-Dichloropropane	9.222	63	25467	10.27	ug/l	98
46) Dibromomethane	9.313	93	14258	10.48	ug/l	98
47) Bromodichloromethane	9.502	83	33444	10.17	ug/l	99
48) Methyl methacrylate	9.301	41	20429	10.64	ug/l	96
49) 1,4-Dioxane	9.301	88	3803	223.92	ug/l	96
51) 4-Methyl-2-Pentanone	10.075	43	116140	53.99	ug/l	100
52) Toluene	10.246	92	63205	10.11	ug/l	98
53) t-1,3-Dichloropropene	10.471	75	37744	10.49	ug/l	98
54) cis-1,3-Dichloropropene	9.935	75	42430	10.42	ug/l	96
55) 1,1,2-Trichloroethane	10.648	97	19260	10.29	ug/l	97
56) Ethyl methacrylate	10.514	69	27430	10.44	ug/l	99
57) 1,3-Dichloropropane	10.794	76	34365	10.36	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.788	63	59736	45.71	ug/l	100
59) 2-Hexanone	10.837	43	76793	53.81	ug/l	99
60) Dibromochloromethane	10.989	129	22938	10.50	ug/l	97
61) 1,2-Dibromoethane	11.093	107	18839	10.39	ug/l	98
64) Tetrachloroethene	10.721	164	27793	10.32	ug/l	99
65) Chlorobenzene	11.520	112	64134	10.29	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.593	131	22628	10.21	ug/l	100
67) Ethyl Benzene	11.599	91	118037	10.22	ug/l	98
68) m/p-Xylenes	11.703	106	87525	20.36	ug/l	100
69) o-Xylene	12.032	106	41062	10.29	ug/l	99
70) Styrene	12.044	104	66365	10.15	ug/l	99
71) Bromoform	12.209	173	13862	10.74	ug/l #	100
73) Isopropylbenzene	12.331	105	111139	10.16	ug/l	99
74) N-amyl acetate	12.148	43	39462	10.71	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.587	83	22079	10.57	ug/l	99
76) 1,2,3-Trichloropropane	12.636	75	31396	19.04	ug/l #	100
77) Bromobenzene	12.611	156	25586	10.27	ug/l	98
78) n-propylbenzene	12.672	91	131325	10.09	ug/l	100
79) 2-Chlorotoluene	12.757	91	71924	10.06	ug/l	99
80) 1,3,5-Trimethylbenzene	12.818	105	91443	10.18	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.386	75	8976	10.66	ug/l	99
82) 4-Chlorotoluene	12.861	91	77021	10.22	ug/l	100
83) tert-Butylbenzene	13.081	119	78009	10.17	ug/l	99
84) 1,2,4-Trimethylbenzene	13.123	105	91848	10.15	ug/l	99
85) sec-Butylbenzene	13.257	105	106990	10.00	ug/l	100
86) p-Isopropyltoluene	13.373	119	97808	10.10	ug/l	99
87) 1,3-Dichlorobenzene	13.367	146	46298	9.99	ug/l	99
88) 1,4-Dichlorobenzene	13.446	146	48210	10.38	ug/l	98
89) n-Butylbenzene	13.696	91	91557	9.93	ug/l	99
90) Hexachloroethane	13.965	117	12091	9.78	ug/l	100
91) 1,2-Dichlorobenzene	13.745	146	41719	10.18	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	4187	10.72	ug/l	94
93) 1,2,4-Trichlorobenzene	15.013	180	27669	10.29	ug/l	100
94) Hexachlorobutadiene	15.117	225	14234	10.12	ug/l	97
95) Naphthalene	15.239	128	61376	10.60	ug/l	100
96) 1,2,3-Trichlorobenzene	15.428	180	24489	10.44	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041221\
Data File : VY004377.D
Acq On : 12 Apr 2021 11:40
Operator : SY/MD
Sample : VSTDICC010
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Quant Time: Apr 13 02:50:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 02:48:44 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041221\
 Data File : VY004377.D
 Acq On : 12 Apr 2021 11:40
 Operator : SY/MD
 Sample : VSTDIC010
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010

Quant Time: Apr 13 02:50:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 02:48:44 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.801 min Scan# 981
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:168 Resp: 211650
Ion Ratio Lower Upper
168 100
99 53.1 45.5 68.3



#2
Dichlorodifluoromethane
Concen: 9.86 ug/l
RT: 1.912 min Scan# 15
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 85 Resp: 19976
Ion Ratio Lower Upper
85 100
87 30.1 16.8 50.4









#3
Chloromethane
Concen: 9.79 ug/l
RT: 2.119 min Scan# 49
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 50 Resp: 29547
Ion Ratio Lower Upper
50 100
52 33.3 26.6 39.8



#4
Vinyl Chloride
Concen: 9.79 ug/l
RT: 2.259 min Scan# 72
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 62 Resp: 31900
Ion Ratio Lower Upper
62 100
64 34.1 25.1 37.7









#5
Bromomethane
Concen: 10.00 ug/l
RT: 2.656 min Scan# 137
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 94 Resp: 15433
Ion Ratio Lower Upper
94 100
96 87.6 72.6 108.8



#6
Chloroethane
Concen: 9.80 ug/l
RT: 2.802 min Scan# 161
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 64 Resp: 18234
Ion Ratio Lower Upper
64 100
66 34.3 26.3 39.5









#7
Trichlorofluoromethane
Concen: 10.08 ug/l
RT: 3.137 min Scan# 216
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:101 Resp: 39272
Ion Ratio Lower Upper
101 100
103 65.1 51.2 76.8



#8
Diethyl Ether
Concen: 10.39 ug/l
RT: 3.546 min Scan# 283
Delta R.T. 0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 74 Resp: 13344
Ion Ratio Lower Upper
74 100
45 104.9 52.3 156.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 9.93 ug/l
RT: 3.911 min Scan# 343
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	101	Resp:	21212
Ion	Ratio	Lower	Upper
101	100		
85	46.2	36.9	55.3
151	79.4	62.0	93.0



#10
Methyl Iodide
Concen: 9.98 ug/l
RT: 4.100 min Scan# 374
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	142	Resp:	25388
Ion	Ratio	Lower	Upper
142	100		
127	45.3	36.2	54.4
141	13.6	11.6	17.4









#11
Tert butyl alcohol
Concen: 53.25 ug/l
RT: 4.972 min Scan# 517
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 59 Resp: 13311
Ion Ratio Lower Upper
59 100
57 10.2 8.1 12.1



#12
1,1-Dichloroethene
Concen: 10.11 ug/l
RT: 3.881 min Scan# 338
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 96 Resp: 21410
Ion Ratio Lower Upper
96 100
61 174.4 132.1 198.1
98 66.5 48.2 72.2









#13
Acrolein
Concen: 57.84 ug/l
RT: 3.741 min Scan# 315
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 56 Resp: 15697
Ion Ratio Lower Upper
56 100
55 69.8 55.4 83.2



#14
Allyl chloride
Concen: 10.18 ug/l
RT: 4.497 min Scan# 439
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40


Tgt Ion: 41 Resp: 40745
Ion Ratio Lower Upper
41 100
39 69.6 56.9 85.3
76 32.8 26.8 40.2








#15
Acrylonitrile
Concen: 53.46 ug/l
RT: 5.186 min Scan# 552
Delta R.T. 0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010







Tgt Ion: 53 Resp: 36836
Ion Ratio Lower Upper
53 100
52 84.0 64.9 97.3
51 37.4 29.3 43.9



#16
Acetone
Concen: 53.29 ug/l
RT: 3.960 min Scan# 351
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 43 Resp: 32702
Ion Ratio Lower Upper
43 100
58 28.3 24.6 37.0









#17
Carbon Disulfide
Concen: 10.17 ug/l
RT: 4.204 min Scan# 391
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 76 Resp: 67514
Ion Ratio Lower Upper
76 100
78 9.9 7.4 11.2



#18
Methyl Acetate
Concen: 10.51 ug/l
RT: 4.491 min Scan# 438
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 43 Resp: 17728
Ion Ratio Lower Upper
43 100
74 21.4 17.2 25.8





#19

Methyl tert-butyl Ether

Concen: 10.34 ug/l

RT: 5.234 min Scan# 560

Delta R.T. -0.006 min

Lab File: VY004377.D





Acq: 12 Apr 2021 11:40

Tgt Ion: 73 Resp: 64543

Ion Ratio Lower Upper

73 100

57 24.8 18.4 27.6



#20

Methylene Chloride

Concen: 10.15 ug/l

RT: 4.728 min Scan# 477

Delta R.T. -0.006 min

Lab File: VY004377.D

Acq: 12 Apr 2021 11:40

Tgt Ion: 84 Resp: 23903




Ion Ratio Lower Upper


84 100

49 125.3 111.2 166.8

51 38.8 33.4 50.2


86 61.4 52.7 79.1








#21
trans-1,2-Dichloroethene
Concen: 9.99 ug/l
RT: 5.241 min Scan# 561
Delta R.T. 0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40


Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010





Tgt Ion: 96 Resp: 23786
Ion Ratio Lower Upper
96 100
61 139.9 116.5 174.7
98 63.7 52.0 78.0




#22
Diisopropyl ether
Concen: 10.16 ug/l
RT: 6.131 min Scan# 707
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40



Tgt Ion: 45 Resp: 84551
Ion Ratio Lower Upper
45 100
43 60.1 46.3 69.5
87 24.3 20.3 30.5
59 11.1 8.6 12.8









#23
 Vinyl Acetate
 Concen: 51.93 ug/l
 RT: 6.076 min Scan# 698
 Delta R.T. 0.000 min
 Lab File: VY004377.D
 Acq: 12 Apr 2021 11:40




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010


Tgt Ion: 43 Resp: 295768
 Ion Ratio Lower Upper
 43 100
 86 9.4 7.6 11.4



#24
 1,1-Dichloroethane
 Concen: 10.13 ug/l
 RT: 6.033 min Scan# 691
 Delta R.T. 0.000 min
 Lab File: VY004377.D
 Acq: 12 Apr 2021 11:40

Tgt Ion: 63 Resp: 43679
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.3 9.9
 100 4.3 2.0 6.0









#25
2-Butanone
Concen: 54.59 ug/l
RT: 7.002 min Scan# 850
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 50226
Ion Ratio Lower Upper
43 100
72 25.6 20.2 30.2



#26
2,2-Dichloropropane
Concen: 10.02 ug/l
RT: 6.996 min Scan# 849
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 77 Resp: 39356
Ion Ratio Lower Upper
77 100
97 21.9 11.1 33.3









#27
 cis-1,2-Dichloroethene
 Concen: 10.22 ug/l
 RT: 7.002 min Scan# 850
 Delta R.T. 0.000 min
 Lab File: VY004377.D
 Acq: 12 Apr 2021 11:40




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010


Tgt Ion: 96 Resp: 27044
 Ion Ratio Lower Upper
 96 100
 61 150.2 0.0 303.2
 98 63.4 0.0 130.0



#28
 Bromochloromethane
 Concen: 9.49 ug/l
 RT: 7.350 min Scan# 907
 Delta R.T. 0.000 min
 Lab File: VY004377.D
 Acq: 12 Apr 2021 11:40

Tgt Ion: 49 Resp: 17751
 Ion Ratio Lower Upper
 49 100
 129 1.6 0.0 3.8
 130 67.7 53.9 80.9





#29

Tetrahydrofuran

Concen: 53.86 ug/l

RT: 7.356 min Scan# 908





Delta R.T. -0.006 min

Lab File: VY004377.D

Acq: 12 Apr 2021 11:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010

Tgt Ion:	42	Resp:	34960
Ion	Ratio	Lower	Upper
42	100		
72	38.3	31.0	46.4
71	36.1	28.8	43.2



#30

Chloroform

Concen: 10.24 ug/l




RT: 7.521 min Scan# 935


Delta R.T. 0.000 min

Lab File: VY004377.D

Acq: 12 Apr 2021 11:40


Tgt Ion:	83	Resp:	43846
Ion	Ratio	Lower	Upper
83	100		
85	63.8	51.3	76.9








#31
Cyclohexane
Concen: 10.05 ug/l
RT: 7.795 min Scan# 980
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40


Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010





Tgt Ion: 56 Resp: 45684
Ion Ratio Lower Upper
56 100
69 45.0 25.8 38.6#
84 95.2 68.2 102.4




#32
1,1,1-Trichloroethane
Concen: 10.09 ug/l
RT: 7.710 min Scan# 966
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40



Tgt Ion: 97 Resp: 39229
Ion Ratio Lower Upper
97 100
99 63.5 51.2 76.8
61 45.3 36.2 54.2





#33

1,2-Dichloroethane-d4

Concen: 9.50 ug/l

RT: 8.149 min Scan# 1038

Delta R.T. -0.006 min

Lab File: VY004377.D

Acq: 12 Apr 2021 11:40

Tgt Ion: 65 Resp: 22461

Ion Ratio Lower Upper

65 100





67 54.0 0.0 106.4

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 8.697 min Scan# 1128

Delta R.T. 0.000 min

Lab File: VY004377.D

Acq: 12 Apr 2021 11:40




Tgt Ion: 114 Resp: 337851


Ion Ratio Lower Upper

114 100

63 20.7 0.0 42.4

88 15.4 0.0 30.8









#35
Dibromofluoromethane
Concen: 9.61 ug/l
RT: 7.728 min Scan# 969
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	113	Resp:	18776
Ion	Ratio	Lower	Upper
113	100		
111	104.4	82.6	123.8
192	20.8	16.1	24.1



#36
1,1-Dichloropropene
Concen: 10.14 ug/l
RT: 7.929 min Scan# 1002
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	75	Resp:	35196
Ion	Ratio	Lower	Upper
75	100		
110	34.9	17.3	51.7
77	32.1	24.3	36.5









#37
Ethyl Acetate
Concen: 10.78 ug/l
RT: 7.088 min Scan# 864
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	43	Resp:	24111
Ion	Ratio	Lower	Upper
43	100		
61	11.5	10.4	15.6
70	9.2	7.5	11.3



#38
Carbon Tetrachloride
Concen: 9.97 ug/l
RT: 7.911 min Scan# 999
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	117	Resp:	28449
Ion	Ratio	Lower	Upper
117	100		
119	102.2	75.8	113.8
121	35.2	25.3	37.9









#39
Methylcyclohexane
Concen: 10.12 ug/l
RT: 9.191 min Scan# 1209
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 83 Resp: 43106
Ion Ratio Lower Upper
83 100
55 70.8 63.0 94.4
98 45.9 39.2 58.8



#40
Benzene
Concen: 10.13 ug/l
RT: 8.167 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 78 Resp: 98654
Ion Ratio Lower Upper
78 100
77 24.7 19.4 29.0









#41
Methacrylonitrile
Concen: 21.87 ug/l
RT: 7.362 min Scan# 909
Delta R.T. 0.018 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	41	Resp:	32625
Ion	Ratio	Lower	Upper
41	100		
39	27.1	0.0	0.0#
67	0.0	0.0	0.0
52	0.0	0.0	0.0



#42
1,2-Dichloroethane
Concen: 10.46 ug/l
RT: 8.246 min Scan# 1054
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	62	Resp:	32021
Ion	Ratio	Lower	Upper
62	100		
98	8.4	0.0	18.2









#43
Isopropyl Acetate
Concen: 10.61 ug/l
RT: 8.283 min Scan# 1060
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 44299
Ion Ratio Lower Upper
43 100
61 25.5 20.9 31.3
87 11.4 9.1 13.7



#44
Trichloroethene
Concen: 10.20 ug/l
RT: 8.947 min Scan# 1169
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 130 Resp: 27271
Ion Ratio Lower Upper
130 100
95 95.6 0.0 190.4





#45

1,2-Dichloropropane

Concen: 10.27 ug/l

RT: 9.222 min Scan# 1214

Delta R.T. -0.006 min

Lab File: VY004377.D

Acq: 12 Apr 2021 11:40

Tgt Ion: 63 Resp: 25467

Ion Ratio Lower Upper

63 100





65 31.5 24.4 36.6

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010



#46

Dibromomethane

Concen: 10.48 ug/l

RT: 9.313 min Scan# 1229

Delta R.T. 0.000 min

Lab File: VY004377.D

Acq: 12 Apr 2021 11:40




Tgt Ion: 93 Resp: 14258


Ion Ratio Lower Upper

93 100

95 79.1 65.8 98.8

174 95.6 77.8 116.6









#47
Bromodichloromethane
Concen: 10.17 ug/l
RT: 9.502 min Scan# 1260
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 83 Resp: 33444
Ion Ratio Lower Upper
83 100
85 62.5 50.7 76.1
127 8.8 6.6 10.0



#48
Methyl methacrylate
Concen: 10.64 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 41 Resp: 20429
Ion Ratio Lower Upper
41 100
69 76.3 62.6 94.0
39 54.2 47.2 70.8









#49
1,4-Dioxane
Concen: 223.92 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 88 Resp: 3803
Ion Ratio Lower Upper
88 100
43 39.4 31.7 47.5
58 74.5 55.7 83.5



#50
Toluene-d8
Concen: 9.63 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 98 Resp: 76051
Ion Ratio Lower Upper
98 100
100 63.7 51.8 77.6









#51
4-Methyl-2-Pentanone
Concen: 53.99 ug/l
RT: 10.075 min Scan# 1354
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 116140
Ion Ratio Lower Upper
43 100
58 36.2 29.2 43.8



#52
Toluene
Concen: 10.11 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 92 Resp: 63205
Ion Ratio Lower Upper
92 100
91 174.4 137.8 206.6









#53
 t-1,3-Dichloropropene
 Concen: 10.49 ug/l
 RT: 10.471 min Scan# 1419
 Delta R.T. 0.000 min
 Lab File: VY004377.D
 Acq: 12 Apr 2021 11:40




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010


Tgt Ion: 75 Resp: 37744
 Ion Ratio Lower Upper
 75 100
 77 30.6 25.2 37.8



#54
 cis-1,3-Dichloropropene
 Concen: 10.42 ug/l
 RT: 9.935 min Scan# 1331
 Delta R.T. 0.000 min
 Lab File: VY004377.D
 Acq: 12 Apr 2021 11:40

Tgt Ion: 75 Resp: 42430
 Ion Ratio Lower Upper
 75 100
 77 29.8 24.6 37.0
 39 49.3 42.1 63.1









#55
1,1,2-Trichloroethane
Concen: 10.29 ug/l
RT: 10.648 min Scan# 1448
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	97	Resp:	19260
Ion	Ratio	Lower	Upper
97	100		
83	86.3	68.4	102.6
85	53.9	46.2	69.4
99	63.5	48.8	73.2



#56
Ethyl methacrylate
Concen: 10.44 ug/l
RT: 10.514 min Scan# 1426
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	69	Resp:	27430
Ion	Ratio	Lower	Upper
69	100		
41	73.3	59.2	88.8
39	38.8	31.0	46.6









#57
1,3-Dichloropropane
Concen: 10.36 ug/l
RT: 10.794 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 76 Resp: 34365
Ion Ratio Lower Upper
76 100
78 32.5 25.9 38.9



#58
2-Chloroethyl Vinyl ether
Concen: 45.71 ug/l
RT: 9.788 min Scan# 1307
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 63 Resp: 59736
Ion Ratio Lower Upper
63 100
106 27.2 21.7 32.5









#59
2-Hexanone
Concen: 53.81 ug/l
RT: 10.837 min Scan# 1479
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 76793
Ion Ratio Lower Upper
43 100
58 51.3 25.4 76.4



#60
Dibromochloromethane
Concen: 10.50 ug/l
RT: 10.989 min Scan# 1504
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 129 Resp: 22938
Ion Ratio Lower Upper
129 100
127 75.4 38.9 116.7









#61
1,2-Dibromoethane
Concen: 10.39 ug/l
RT: 11.093 min Scan# 1521
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:107 Resp: 18839
Ion Ratio Lower Upper
107 100
109 92.7 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 9.64 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 95 Resp: 25626
Ion Ratio Lower Upper
95 100
174 81.3 0.0 162.4
176 80.6 0.0 156.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.495 min Scan# 1587
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:117 Resp: 297999
Ion Ratio Lower Upper
117 100
82 55.0 44.9 67.3
119 32.1 25.4 38.0



#64
Tetrachloroethene
Concen: 10.32 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:164 Resp: 27793
Ion Ratio Lower Upper
164 100
166 126.9 102.1 153.1
129 91.5 74.8 112.2
131 88.9 72.2 108.4









#65
Chlorobenzene
Concen: 10.29 ug/l
RT: 11.520 min Scan# 1591
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:112 Resp: 64134
Ion Ratio Lower Upper
112 100
114 32.6 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 10.21 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:131 Resp: 22628
Ion Ratio Lower Upper
131 100
133 95.3 47.6 142.9
119 65.9 32.6 98.0









#67
Ethyl Benzene
Concen: 10.22 ug/l
RT: 11.599 min Scan# 1604
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 91 Resp: 118037
Ion Ratio Lower Upper
91 100
106 32.6 25.1 37.7



#68
m/p-Xylenes
Concen: 20.36 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 106 Resp: 87525
Ion Ratio Lower Upper
106 100
91 203.7 163.2 244.8









#69
o-Xylene
Concen: 10.29 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:106 Resp: 41062
Ion Ratio Lower Upper
106 100
91 212.7 107.4 322.2



#70
Styrene
Concen: 10.15 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:104 Resp: 66365
Ion Ratio Lower Upper
104 100
78 50.1 39.9 59.9
103 55.6 43.4 65.2









#71
Bromoform
Concen: 10.74 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:173 Resp: 13862
Ion Ratio Lower Upper
173 100
175 48.3 24.3 72.8
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:152 Resp: 138219
Ion Ratio Lower Upper
152 100
115 58.7 29.6 88.8
150 161.1 0.0 348.4





#73

Isopropylbenzene

Concen: 10.16 ug/l

RT: 12.331 min Scan# 1724

Delta R.T. 0.000 min

Lab File: VY004377.D





Acq: 12 Apr 2021 11:40

Tgt Ion:105 Resp: 111139

Ion Ratio Lower Upper

105 100

120 26.7 13.2 39.5



#74

N-ethyl acetate

Concen: 10.71 ug/l

RT: 12.148 min Scan# 1694

Delta R.T. 0.000 min

Lab File: VY004377.D

Acq: 12 Apr 2021 11:40

Tgt Ion: 43 Resp: 39462




Ion Ratio Lower Upper


43 100

70 36.1 30.4 45.6

55 21.6 17.4 26.0

61 21.2 18.2 27.2









#75
1,1,2,2-Tetrachloroethane
Concen: 10.57 ug/l
RT: 12.587 min Scan# 1766
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	83	Resp:	22079
Ion	Ratio	Lower	Upper
83	100		
131	10.5	5.1	15.4
85	62.5	31.6	94.8



#76
1,2,3-Trichloropropane
Concen: 19.04 ug/l
RT: 12.636 min Scan# 1774
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	75	Resp:	31396
Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0









#77
Bromobenzene
Concen: 10.27 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	156	Resp:	25586
Ion Ratio	Lower	Upper	
156	100		
77	193.5	98.6	295.7
158	95.9	48.4	145.2



#78
n-propylbenzene
Concen: 10.09 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	91	Resp:	131325
Ion Ratio	Lower	Upper	
91	100		
120	23.0	11.5	34.4









#79
2-Chlorotoluene
Concen: 10.06 ug/l
RT: 12.757 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 91 Resp: 71924
Ion Ratio Lower Upper
91 100
126 34.0 16.7 50.1



#80
1,3,5-Trimethylbenzene
Concen: 10.18 ug/l
RT: 12.818 min Scan# 1804
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 105 Resp: 91443
Ion Ratio Lower Upper
105 100
120 49.1 24.6 74.0









#81
trans-1,4-Dichloro-2-butene
Concen: 10.66 ug/l
RT: 12.386 min Scan# 1733
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion: 75 Resp: 8976
Ion Ratio Lower Upper
75 100
53 93.1 75.0 112.6
89 46.3 36.1 54.1



#82
4-Chlorotoluene
Concen: 10.22 ug/l
RT: 12.861 min Scan# 1811
Delta R.T. 0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion: 91 Resp: 77021
Ion Ratio Lower Upper
91 100
126 34.1 16.9 50.7









#83
tert-Butylbenzene
Concen: 10.17 ug/l
RT: 13.081 min Scan# 1847
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	119	Resp:	78009
Ion Ratio	Lower	Upper	
119	100		
91	64.6	32.6	98.0
134	26.1	12.9	38.6



#84
1,2,4-Trimethylbenzene
Concen: 10.15 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	105	Resp:	91848
Ion Ratio	Lower	Upper	
105	100		
120	44.5	22.4	67.3









#85
 sec-Butylbenzene
 Concen: 10.00 ug/l
 RT: 13.257 min Scan# 1876
 Delta R.T. 0.000 min
 Lab File: VY004377.D
 Acq: 12 Apr 2021 11:40




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010


Tgt Ion:105 Resp: 106990
 Ion Ratio Lower Upper
 105 100
 134 20.4 10.1 30.3



#86
 p-Isopropyltoluene
 Concen: 10.10 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.000 min
 Lab File: VY004377.D
 Acq: 12 Apr 2021 11:40

Tgt Ion:119 Resp: 97808
 Ion Ratio Lower Upper
 119 100
 134 26.0 13.3 39.8
 91 24.0 11.9 35.7









#87
1,3-Dichlorobenzene
Concen: 9.99 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:146 Resp: 46298
Ion Ratio Lower Upper
146 100
111 41.2 20.4 61.2
148 65.0 32.1 96.5



#88
1,4-Dichlorobenzene
Concen: 10.38 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:146 Resp: 48210
Ion Ratio Lower Upper
146 100
111 41.7 19.6 58.8
148 63.3 31.9 95.5









#89
n-Butylbenzene
Concen: 9.93 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	91	Resp:	91557
Ion	Ratio	Lower	Upper
91	100		
92	52.9	26.6	79.8
134	26.4	12.7	38.1



#90
Hexachloroethane
Concen: 9.78 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	117	Resp:	12091
Ion	Ratio	Lower	Upper
117	100		
201	73.2	36.6	109.9









#91
1,2-Dichlorobenzene
Concen: 10.18 ug/l
RT: 13.745 min Scan# 1956
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	146	Resp:	41719
Ion Ratio	Lower	Upper	
146	100		
111	42.4	20.9	62.7
148	64.1	31.9	95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 10.72 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.006 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	75	Resp:	4187
Ion Ratio	Lower	Upper	
75	100		
155	77.1	40.6	121.8
157	97.8	53.0	159.0









#93
1,2,4-Trichlorobenzene
Concen: 10.29 ug/l
RT: 15.013 min Scan# 2164
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010


Tgt Ion:	180	Resp:	27669
Ion Ratio	Lower	Upper	
180	100		
182	94.4	47.3	141.8
145	31.3	15.5	46.5



#94
Hexachlorobutadiene
Concen: 10.12 ug/l
RT: 15.117 min Scan# 2181
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:	225	Resp:	14234
Ion Ratio	Lower	Upper	
225	100		
223	61.1	31.3	93.9
227	61.1	32.2	96.6









#95
Naphthalene
Concen: 10.60 ug/l
RT: 15.239 min Scan# 2201
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010

Tgt Ion:128 Resp: 61376
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 10.44 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.000 min
Lab File: VY004377.D
Acq: 12 Apr 2021 11:40

Tgt Ion:180 Resp: 24489
Ion Ratio Lower Upper
180 100
182 95.9 47.6 142.8
145 32.6 16.3 48.8

