



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041224\
Data File : VY017918.D
Acq On : 12 Apr 2024 08:55
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Title : SW846 8260
Last Update : Sat Mar 30 05:58:13 2024



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	14482	PASS
75	95	30	60	54.4	38240	PASS
95	95	100	100	100.0	70320	PASS
96	95	5	9	6.6	4612	PASS
173	174	0.00	2	1.6	1122	PASS
174	95	50	100	96.9	68139	PASS
175	174	5	9	7.3	4956	PASS
176	174	95	101	95.7	65243	PASS
177	176	5	9	6.5	4272	PASS