

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041323\
 Data File : VY013297.D
 Acq On : 13 Apr 2023 15:42
 Operator : KP/MD
 Sample : 02224-09
 Misc : 4.95g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU3-FC-BACKFILL-05

Quant Time: Apr 14 03:00:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	167354	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.685	114	294951	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	281999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	131603	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	123784	61.547	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.100%
35) Dibromofluoromethane	7.710	113	103431	54.167	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.340%
50) Toluene-d8	10.173	98	360212	50.681	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.360%
62) 4-Bromofluorobenzene	12.477	95	117575	50.913	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041323\
Data File : VY013297.D
Acq On : 13 Apr 2023 15:42
Operator : KP/MD
Sample : 02224-09
Misc : 4.95g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU3-FC-BACKFILL-05

Quant Time: Apr 14 03:00:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 05:55:16 2023
Response via : Initial Calibration

