

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041323\
 Data File : VY013298.D
 Acq On : 13 Apr 2023 16:05
 Operator : KP/MD
 Sample : 02224-11
 Misc : 4.73g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU3-FC-BACKFILL-06

Quant Time: Apr 14 03:00:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 05:55:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	167071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	292203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.484	117	277173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	131040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	123428	61.474	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.940%	
35) Dibromofluoromethane	7.710	113	104442	55.211	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.420%	
50) Toluene-d8	10.173	98	365507	51.909	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.820%	
62) 4-Bromofluorobenzene	12.477	95	113031	49.406	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.820%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041323\
Data File : VY013298.D
Acq On : 13 Apr 2023 16:05
Operator : KP/MD
Sample : 02224-11
Misc : 4.73g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU3-FC-BACKFILL-06

Quant Time: Apr 14 03:00:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041223S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 05:55:16 2023
Response via : Initial Calibration

