

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041521\
 Data File : VY004436.D
 Acq On : 15 Apr 2021 13:01
 Operator : SY/MD
 Sample : M2047-06
 Misc : 7.85G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Apr 16 05:03:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	137845	50.00	ug/l	# 0.02
34) 1,4-Difluorobenzene	8.709	114	253767	50.00	ug/l	# 0.01
63) Chlorobenzene-d5	11.508	117	254596	50.00	ug/l	# 0.01
72) 1,4-Dichlorobenzene-d4	13.434	152	94582	50.00	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.118	65	283981	185.42	ug/l	-0.04
Spiked Amount 50.000	Range 50 - 163		Recovery = 370.84%#			
35) Dibromofluoromethane	7.734	113	17213	12.22	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 24.44%#			
50) Toluene-d8	10.227	98	5297951	884.42	ug/l	0.04
Spiked Amount 50.000	Range 49 - 140		Recovery = 1768.84%#			
62) 4-Bromofluorobenzene	12.489	95	1049643	516.36	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 1032.72%#			
Target Compounds						
						Qvalue
3) Chloromethane	2.077	50	28322	15.34	ug/l	# 67
4) Vinyl Chloride	2.186	62	8040	4.12	ug/l	# 43
5) Bromomethane	2.656	94	1076	1.13	ug/l	85
6) Chloroethane	2.790	64	1328	1.14	ug/l	91
10) Methyl Iodide	4.101	142	6096	3.76	ug/l	100
11) Tert butyl alcohol	4.796	59	5387	35.69	ug/l	# 73
13) Acrolein	3.704	56	565031	3367.53	ug/l	# 1
14) Allyl chloride	4.600	41	995669	389.07	ug/l	# 71
15) Acrylonitrile	5.265	53	2067105	4673.27	ug/l	# 42
16) Acetone	3.960	43	10055868	33492.35	ug/l	# 54
17) Carbon Disulfide	4.198	76	14998	3.64	ug/l	# 94
18) Methyl Acetate	4.722	43	25392400	23803.23	ug/l	# 54
19) Methyl tert-butyl Ether	5.271	73	12187	3.03	ug/l	# 1
20) Methylene Chloride	4.613	84	370856	250.71	ug/l	# 11
23) Vinyl Acetate	6.131	43	952315	258.62	ug/l	# 83
24) 1,1-Dichloroethane	6.131	63	378307	136.90	ug/l	# 83
25) 2-Butanone	7.027	43	2337219	4006.19	ug/l	# 50
26) 2,2-Dichloropropane	6.862	77	810073	326.69	ug/l	# 54
28) Bromochloromethane	7.368	49	5899	4.70	ug/l	# 18
29) Tetrahydrofuran	7.380	42	52305	125.49	ug/l	# 38
30) Chloroform	7.521	83	7140079	2629.30	ug/l	# 32
31) Cyclohexane	7.844	56	27174106	9678.46	ug/l	# 88
32) 1,1,1-Trichloroethane	7.710	97	11762	4.77	ug/l	# 1
36) 1,1-Dichloropropene	8.051	75	3390	1.34	ug/l	# 1
37) Ethyl Acetate	7.027	43	2337219	1439.31	ug/l	# 70
38) Carbon Tetrachloride	7.801	117	12334	5.66	ug/l	# 16
39) Methylcyclohexane	9.252	83	46054495	14834.67	ug/l	# 86
40) Benzene	8.197	78	263326	37.16	ug/l	# 78
41) Methacrylonitrile	7.387	41	1356930	1378.87	ug/l	# 100
42) 1,2-Dichloroethane	8.338	62	234155	105.34	ug/l	# 1
43) Isopropyl Acetate	8.295	43	3664316	1190.34	ug/l	# 55
45) 1,2-Dichloropropane	9.197	63	822309	449.61	ug/l	# 1
46) Dibromomethane	9.301	93	132729	133.30	ug/l	# 1
47) Bromodichloromethane	9.520	83	4695018	1947.61	ug/l	# 41
48) Methyl methacrylate	9.411	41	8534102	6019.93	ug/l	# 27
49) 1,4-Dioxane	9.319	88	344	26.45	ug/l	# 22

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041521\
 Data File : VY004436.D
 Acq On : 15 Apr 2021 13:01
 Operator : SY/MD
 Sample : M2047-06
 Misc : 7.85G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Apr 16 05:03:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

51) 4-Methyl-2-Pentanone	10.020	43	30119406	19047.62	ug/l #	46
52) Toluene	10.276	92	496983	108.38	ug/l #	58
54) cis-1,3-Dichloropropene	9.904	75	7221	2.39	ug/l #	1
55) 1,1,2-Trichloroethane	10.660	97	16663930	12145.28	ug/l #	22
56) Ethyl methacrylate	10.538	69	6714671	3374.86	ug/l #	21
57) 1,3-Dichloropropane	10.807	76	5452	2.23	ug/l #	1
58) 2-Chloroethyl Vinyl ether	9.746	63	779827	766.04	ug/l #	48
59) 2-Hexanone	10.843	43	14422764	13606.09	ug/l #	32
60) Dibromochloromethane	10.953	129	87087	53.41	ug/l #	10
61) 1,2-Dibromoethane	11.099	107	47252	35.72	ug/l #	1
65) Chlorobenzene	11.441	112	341409	65.71	ug/l #	54
67) Ethyl Benzene	11.599	91	3899335	401.11	ug/l #	26
68) m/p-Xylenes	11.733	106	17625499	4879.34	ug/l #	64
69) o-Xylene	12.050	106	5259860	1553.84	ug/l	98
70) Styrene	12.050	104	265634	47.38	ug/l #	1
73) Isopropylbenzene	12.343	105	20344608	2783.63	ug/l	80
74) N-amyl acetate	12.148	43	5712034	2321.41	ug/l #	46
75) 1,1,2,2-Tetrachloroethane	12.544	83	168599	121.93	ug/l #	1
76) 1,2,3-Trichloropropane	12.690	75	412553	406.40	ug/l #	100
77) Bromobenzene	12.593	156	5997	3.66	ug/l #	1
78) n-propylbenzene	12.782	91	4699613	544.31	ug/l #	1
79) 2-Chlorotoluene	12.782	91	4699613	995.28	ug/l #	41
80) 1,3,5-Trimethylbenzene	12.825	105	9680241	1629.56	ug/l	95
81) trans-1,4-Dichloro-2-b...	12.349	75	379655	669.04	ug/l #	1
82) 4-Chlorotoluene	12.825	91	1260497	252.79	ug/l #	48
83) tert-Butylbenzene	13.087	119	494877	97.33	ug/l	86
84) 1,2,4-Trimethylbenzene	13.264	105	4188852	698.75	ug/l #	31
85) sec-Butylbenzene	13.264	105	4188852	590.60	ug/l #	54
86) p-Isopropyltoluene	13.373	119	1846693	286.23	ug/l	96
89) n-Butylbenzene	13.703	91	4792742	778.28	ug/l #	74
90) Hexachloroethane	13.977	117	870639	1005.83	ug/l #	13
92) 1,2-Dibromo-3-Chloropr...	14.379	75	9996	40.29	ug/l #	10
95) Naphthalene	15.239	128	2272849	582.18	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 16 05:03:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.819 min Scan# 984
Delta R.T. 0.018 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:168 Resp: 137845
Ion Ratio Lower Upper
168 100
99 80.2 45.5 68.3#



#3
Chloromethane
Concen: 15.34 ug/l
RT: 2.077 min Scan# 42
Delta R.T. -0.042 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 50 Resp: 28322
Ion Ratio Lower Upper
50 100
52 14.7 26.6 39.8#









#4
 Vinyl Chloride
 Concen: 4.12 ug/l
 RT: 2.186 min Scan# 60
 Delta R.T. -0.073 min
 Lab File: VY004436.D
 Acq: 15 Apr 2021 13:01




Instrument :
 MSVOA_Y
 ClientSampled :


Tgt Ion: 62 Resp: 8040
 Ion Ratio Lower Upper
 62 100
 64 0.0 25.1 37.7#



#5
 Bromomethane
 Concen: 1.13 ug/l
 RT: 2.656 min Scan# 137
 Delta R.T. 0.000 min
 Lab File: VY004436.D
 Acq: 15 Apr 2021 13:01

Tgt Ion: 94 Resp: 1076
 Ion Ratio Lower Upper
 94 100
 96 76.7 72.6 108.8









#6
Chloroethane
Concen: 1.14 ug/l
RT: 2.790 min Scan# 159
Delta R.T. -0.012 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 64 Resp: 1328
Ion Ratio Lower Upper
64 100
66 38.2 26.3 39.5



#10
Methyl Iodide
Concen: 3.76 ug/l
RT: 4.101 min Scan# 374
Delta R.T. -0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 142 Resp: 6096
Ion Ratio Lower Upper
142 100
127 45.3 36.2 54.4
141 14.1 11.6 17.4









#11
Tert butyl alcohol
Concen: 35.69 ug/l
RT: 4.796 min Scan# 488
Delta R.T. -0.183 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 59 Resp: 5387
Ion Ratio Lower Upper
59 100
57 0.0 8.1 12.1#



#13
Acrolein
Concen: 3367.53 ug/l
RT: 3.704 min Scan# 309
Delta R.T. -0.036 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 56 Resp: 565031
Ion Ratio Lower Upper
56 100
55 2185.0 55.4 83.2#





#14

Allyl chloride

Concen: 389.07 ug/l

RT: 4.600 min Scan# 456





Delta R.T. 0.104 min

Lab File: VY004436.D

Acq: 15 Apr 2021 13:01

Instrument :
MSVOA_Y
ClientSampled :

Tgt Ion: 41 Resp: 995669
Ion Ratio Lower Upper
41 100
39 59.0 56.9 85.3
76 0.0 26.8 40.2#



#15

Acrylonitrile

Concen: 4673.27 ug/l




RT: 5.265 min Scan# 565


Delta R.T. 0.086 min

Lab File: VY004436.D

Acq: 15 Apr 2021 13:01

Tgt Ion: 53 Resp: 2067105
Ion Ratio Lower Upper
53 100
52 20.2 64.9 97.3#
51 56.6 29.3 43.9#









#16
Acetone
Concen: 33492.35 ug/l
RT: 3.960 min Scan# 351
Delta R.T. 0.000 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp:10055868
Ion Ratio Lower Upper
43 100
58 5.4 24.6 37.0#



#17
Carbon Disulfide
Concen: 3.64 ug/l
RT: 4.198 min Scan# 390
Delta R.T. -0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 76 Resp: 14998
Ion Ratio Lower Upper
76 100
78 7.0 7.4 11.2#









#18
Methyl Acetate
Concen: 23803.23 ug/l
RT: 4.722 min Scan# 476
Delta R.T. 0.232 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 43 Resp:25392400
Ion Ratio Lower Upper
43 100
74 0.0 17.2 25.8#



#19
Methyl tert-butyl Ether
Concen: 3.03 ug/l
RT: 5.271 min Scan# 566
Delta R.T. 0.031 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01


Tgt Ion: 73 Resp: 12187
Ion Ratio Lower Upper
73 100
57 250090.8 18.4 27.6#








#20
Methylene Chloride
Concen: 250.71 ug/l
RT: 4.613 min Scan# 458
Delta R.T. -0.122 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Instrument :
MSVOA_Y
ClientSampled :







Tgt Ion:	84	Resp:	370856
Ion	Ratio	Lower	Upper
84	100		
49	3.3	111.2	166.8#
51	27.6	33.4	50.2#
86	0.3	52.7	79.1#



#23
Vinyl Acetate
Concen: 258.62 ug/l
RT: 6.131 min Scan# 707
Delta R.T. 0.055 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01


Tgt Ion:	43	Resp:	952315
Ion	Ratio	Lower	Upper
43	100		
86	3.5	7.6	11.4#








#24
1,1-Dichloroethane
Concen: 136.90 ug/l
RT: 6.131 min Scan# 707
Delta R.T. 0.098 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Instrument :
MSVOA_Y
ClientSampled :







Tgt Ion: 63 Resp: 378307
Ion Ratio Lower Upper
63 100
98 0.0 3.3 9.9#
100 0.0 2.0 6.0#



#25
2-Butanone
Concen: 4006.19 ug/l
RT: 7.027 min Scan# 854
Delta R.T. 0.025 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 43 Resp: 2337219
Ion Ratio Lower Upper
43 100
72 0.2 20.2 30.2#









#26
2,2-Dichloropropane
Concen: 326.69 ug/l
RT: 6.862 min Scan# 827
Delta R.T. -0.134 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 77 Resp: 810073
Ion Ratio Lower Upper
77 100
97 0.0 11.1 33.3#



#28
Bromochloromethane
Concen: 4.70 ug/l
RT: 7.368 min Scan# 910
Delta R.T. 0.019 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 49 Resp: 5899
Ion Ratio Lower Upper
49 100
129 0.0 0.0 3.8
130 0.0 53.9 80.9#





#29

Tetrahydrofuran

Concen: 125.49 ug/l

RT: 7.380 min Scan# 912





Delta R.T. 0.018 min

Lab File: VY004436.D

Acq: 15 Apr 2021 13:01

Instrument :
MSVOA_Y
ClientSampled :

Tgt Ion:	42	Resp:	52305
Ion	Ratio	Lower	Upper
42	100		
72	0.0	31.0	46.4#
71	0.0	28.8	43.2#



#30

Chloroform

Concen: 2629.30 ug/l




RT: 7.521 min Scan# 935


Delta R.T. 0.000 min

Lab File: VY004436.D

Acq: 15 Apr 2021 13:01


Tgt Ion:	83	Resp:	7140079
Ion	Ratio	Lower	Upper
83	100		
85	10.9	51.3	76.9#








#31
Cyclohexane
Concen: 9678.46 ug/l
RT: 7.844 min Scan# 988
Delta R.T. 0.049 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Instrument :
MSVOA_Y
ClientSampled :







Tgt Ion: 56 Resp: 27174106
Ion Ratio Lower Upper
56 100
69 18.7 25.8 38.6#
84 92.2 68.2 102.4



#32
1,1,1-Trichloroethane
Concen: 4.77 ug/l
RT: 7.710 min Scan# 966
Delta R.T. -0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 97 Resp: 11762
Ion Ratio Lower Upper
97 100
99 402.8 51.2 76.8#
61 0.0 36.2 54.2#









#33
1,2-Dichloroethane-d4
Concen: 185.42 ug/l
RT: 8.118 min Scan# 1033
Delta R.T. -0.036 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 65 Resp: 283981
Ion Ratio Lower Upper
65 100
67 344.2 0.0 106.4#



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.709 min Scan# 1130
Delta R.T. 0.012 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 114 Resp: 253767
Ion Ratio Lower Upper
114 100
63 117.8 0.0 42.4#
88 17.3 0.0 30.8









#35
Dibromofluoromethane
Concen: 12.22 ug/l
RT: 7.734 min Scan# 970
Delta R.T. 0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:	113	Resp:	17213
Ion Ratio	Lower	Upper	
113	100		
111	107.5	82.6	123.8
192	20.1	16.1	24.1



#36
1,1-Dichloropropene
Concen: 1.34 ug/l
RT: 8.051 min Scan# 1022
Delta R.T. 0.122 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion:	75	Resp:	3390
Ion Ratio	Lower	Upper	
75	100		
110	1.3	17.3	51.7#
77	729.2	24.3	36.5#





#37

Ethyl Acetate

Concen: 1439.31 ug/l

RT: 7.027 min Scan# 854

Delta R.T. -0.061 min

Lab File: VY004436.D

Acq: 15 Apr 2021 13:01





Tgt Ion: 43 Resp: 2337219

Ion Ratio Lower Upper

43 100

61 0.0 10.4 15.6#

70 0.0 7.5 11.3#



#38

Carbon Tetrachloride

Concen: 5.66 ug/l

RT: 7.801 min Scan# 981

Delta R.T. -0.110 min

Lab File: VY004436.D

Acq: 15 Apr 2021 13:01




Tgt Ion: 117 Resp: 12334


Ion Ratio Lower Upper

117 100

119 3.3 75.8 113.8#

121 1.5 25.3 37.9#





#39

Methylcyclohexane

Concen: 14834.67 ug/l

RT: 9.252 min Scan# 1219

Delta R.T. 0.061 min





Lab File: VY004436.D

Acq: 15 Apr 2021 13:01

Instrument :
MSVOA_Y
ClientSampled :

Tgt Ion: 83 Resp:46054495

Ion	Ratio	Lower	Upper
83	100		
55	73.2	63.0	94.4
98	67.8	39.2	58.8



#40

Benzene

Concen: 37.16 ug/l

RT: 8.197 min Scan# 1046




Delta R.T. 0.025 min


Lab File: VY004436.D

Acq: 15 Apr 2021 13:01

Tgt Ion: 78 Resp: 263326


Ion	Ratio	Lower	Upper
78	100		
77	13.3	19.4	29.0








#41
Methacrylonitrile
Concen: 1378.87 ug/l
RT: 7.387 min Scan# 913
Delta R.T. 0.043 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion: 41 Resp: 1356930
Ion Ratio Lower Upper
41 100
39 60.9 0.0 0.0#
67 0.0 0.0 0.0
52 4.5 0.0 0.0#




#42
1,2-Dichloroethane
Concen: 105.34 ug/l
RT: 8.338 min Scan# 1069
Delta R.T. 0.092 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01



Tgt Ion: 62 Resp: 234155
Ion Ratio Lower Upper
62 100
98 1274.7 0.0 18.2#









#43

Isopropyl Acetate
Concen: 1190.34 ug/l
RT: 8.295 min Scan# 1062
Delta R.T. 0.012 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Instrument :
MSVOA_Y
ClientSampled :




Tgt Ion: 43 Resp: 3664316
Ion Ratio Lower Upper
43 100
61 0.0 20.9 31.3#
87 0.0 9.1 13.7#




#45

1,2-Dichloropropane
Concen: 449.61 ug/l
RT: 9.197 min Scan# 1210
Delta R.T. -0.030 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 63 Resp: 822309
Ion Ratio Lower Upper
63 100
65 225.3 24.4 36.6#





#46

Dibromomethane

Concen: 133.30 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. -0.012 min

Lab File: VY004436.D

Acq: 15 Apr 2021 13:01

Tgt Ion: 93 Resp: 132729

Ion Ratio Lower Upper

93 100





95 1738.1 65.8 98.8#

174 0.0 77.8 116.6#

Instrument :

MSVOA_Y

ClientSampled :



#47

Bromodichloromethane

Concen: 1947.61 ug/l

RT: 9.520 min Scan# 1263

Delta R.T. 0.018 min

Lab File: VY004436.D

Acq: 15 Apr 2021 13:01




Tgt Ion: 83 Resp: 4695018


Ion Ratio Lower Upper

83 100

85 13.5 50.7 76.1#

127 0.0 6.6 10.0#









#48
Methyl methacrylate
Concen: 6019.93 ug/l
RT: 9.411 min Scan# 1245
Delta R.T. 0.110 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 41 Resp: 8534102
Ion Ratio Lower Upper
41 100
69 189.7 62.6 94.0#
39 60.0 47.2 70.8



#49
1,4-Dioxane
Concen: 26.45 ug/l
RT: 9.319 min Scan# 1230
Delta R.T. 0.012 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 88 Resp: 344
Ion Ratio Lower Upper
88 100
43 0.0 31.7 47.5#
58 0.0 55.7 83.5#









#50
Toluene-d8
Concen: 884.42 ug/l
RT: 10.227 min Scan# 1379
Delta R.T. 0.043 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 98 Resp: 5297951
Ion Ratio Lower Upper
98 100
100 3.5 51.8 77.6#



#51
4-Methyl-2-Pentanone
Concen: 19047.62 ug/l
RT: 10.020 min Scan# 1345
Delta R.T. -0.055 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 43 Resp: 30119406
Ion Ratio Lower Upper
43 100
58 4.6 29.2 43.8#









#52
Toluene
Concen: 108.38 ug/l
RT: 10.276 min Scan# 1387
Delta R.T. 0.025 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 92 Resp: 496983
Ion Ratio Lower Upper
92 100
91 229.6 137.8 206.6#



#54
cis-1,3-Dichloropropene
Concen: 2.39 ug/l
RT: 9.904 min Scan# 1326
Delta R.T. -0.030 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01


Tgt Ion: 75 Resp: 7221
Ion Ratio Lower Upper
75 100
77 2384.5 24.6 37.0#
39 37950.9 42.1 63.1#








#55
1,1,2-Trichloroethane
Concen: 12145.28 ug/l
RT: 10.660 min Scan# 1450
Delta R.T. 0.012 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Instrument :
MSVOA_Y
ClientSampled :







Tgt Ion: 97 Resp:16663930
Ion Ratio Lower Upper
97 100
83 7.7 68.4 102.6#
85 8.9 46.2 69.4#
99 0.3 48.8 73.2#



#56
Ethyl methacrylate
Concen: 3374.86 ug/l
RT: 10.538 min Scan# 1430
Delta R.T. 0.025 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 69 Resp: 6714671
Ion Ratio Lower Upper
69 100
41 0.0 59.2 88.8#
39 0.0 31.0 46.6#









#57
1,3-Dichloropropane
Concen: 2.23 ug/l
RT: 10.807 min Scan# 1474
Delta R.T. 0.012 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 76 Resp: 5452
Ion Ratio Lower Upper
76 100
78 1230.1 25.9 38.9#



#58
2-Chloroethyl Vinyl ether
Concen: 766.04 ug/l
RT: 9.746 min Scan# 1300
Delta R.T. -0.042 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 63 Resp: 779827
Ion Ratio Lower Upper
63 100
106 0.0 21.7 32.5#









#59
2-Hexanone
Concen: 13606.09 ug/l
RT: 10.843 min Scan# 1480
Delta R.T. 0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 43 Resp:14422764
Ion Ratio Lower Upper
43 100
58 3.9 25.4 76.4#



#60
Dibromochloromethane
Concen: 53.41 ug/l
RT: 10.953 min Scan# 1498
Delta R.T. -0.036 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion:129 Resp: 87087
Ion Ratio Lower Upper
129 100
127 0.0 38.9 116.7#









#61
1,2-Dibromoethane
Concen: 35.72 ug/l
RT: 11.099 min Scan# 1522
Delta R.T. 0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:107 Resp: 47252
Ion Ratio Lower Upper
107 100
109 4169.8 75.5 113.3#



#62
4-Bromofluorobenzene
Concen: 516.36 ug/l
RT: 12.489 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 95 Resp: 1049643
Ion Ratio Lower Upper
95 100
174 7.5 0.0 162.4
176 7.3 0.0 156.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.508 min Scan# 1589
Delta R.T. 0.012 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:117 Resp: 254596
Ion Ratio Lower Upper
117 100
82 312.4 44.9 67.3#
119 32.4 25.4 38.0



#65
Chlorobenzene
Concen: 65.71 ug/l
RT: 11.441 min Scan# 1578
Delta R.T. -0.079 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion:112 Resp: 341409
Ion Ratio Lower Upper
112 100
114 6.4 25.8 38.6#









#67
Ethyl Benzene
Concen: 401.11 ug/l
RT: 11.599 min Scan# 1604
Delta R.T. 0.000 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 91 Resp: 3899335
Ion Ratio Lower Upper
91 100
106 72.3 25.1 37.7#



#68
m/p-Xylenes
Concen: 4879.34 ug/l
RT: 11.733 min Scan# 1626
Delta R.T. 0.025 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion:106 Resp:17625499
Ion Ratio Lower Upper
106 100
91 148.8 163.2 244.8#









#69
o-Xylene
Concen: 1553.84 ug/l
RT: 12.050 min Scan# 1678
Delta R.T. 0.018 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:106 Resp: 5259860
Ion Ratio Lower Upper
106 100
91 211.7 107.4 322.2



#70
Styrene
Concen: 47.38 ug/l
RT: 12.050 min Scan# 1678
Delta R.T. 0.000 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion:104 Resp: 265634
Ion Ratio Lower Upper
104 100
78 356.3 39.9 59.9#
103 275.7 43.4 65.2#









#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.434 min Scan# 1905
Delta R.T. 0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:152 Resp: 94582
Ion Ratio Lower Upper
152 100
115 0.0 29.6 88.8#
150 53.4 0.0 348.4



#73
Isopropylbenzene
Concen: 2783.63 ug/l
RT: 12.343 min Scan# 1726
Delta R.T. 0.012 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion:105 Resp:20344608
Ion Ratio Lower Upper
105 100
120 36.7 13.2 39.5









#74
N-amy1 acetate
Concen: 2321.41 ug/l
RT: 12.148 min Scan# 1694
Delta R.T. 0.000 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 43 Resp: 5712034
Ion Ratio Lower Upper
43 100
70 0.0 30.4 45.6#
55 0.0 17.4 26.0#
61 0.0 18.2 27.2#



#75
1,1,2,2-Tetrachloroethane
Concen: 121.93 ug/l
RT: 12.544 min Scan# 1759
Delta R.T. -0.042 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 83 Resp: 168599
Ion Ratio Lower Upper
83 100
131 0.0 5.1 15.4#
85 234.4 31.6 94.8#









#76
1,2,3-Trichloropropane
Concen: 406.40 ug/l
RT: 12.690 min Scan# 1783
Delta R.T. 0.055 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 75 Resp: 412553
Ion Ratio Lower Upper
75 100
77 636.7 0.0 0.0#



#77
Bromobenzene
Concen: 3.66 ug/l
RT: 12.593 min Scan# 1767
Delta R.T. -0.018 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 156 Resp: 5997
Ion Ratio Lower Upper
156 100
77 3692.7 98.6 295.7#
158 0.0 48.4 145.2#









#78
n-propylbenzene
Concen: 544.31 ug/l
RT: 12.782 min Scan# 1798
Delta R.T. 0.110 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 91 Resp: 4699613
Ion Ratio Lower Upper
91 100
120 260.8 11.5 34.4#



#79
2-Chlorotoluene
Concen: 995.28 ug/l
RT: 12.782 min Scan# 1798
Delta R.T. 0.025 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 91 Resp: 4699613
Ion Ratio Lower Upper
91 100
126 0.0 16.7 50.1#









#80
1,3,5-Trimethylbenzene
Concen: 1629.56 ug/l
RT: 12.825 min Scan# 1805
Delta R.T. 0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:105 Resp: 9680241
Ion Ratio Lower Upper
105 100
120 52.5 24.6 74.0



#81
trans-1,4-Dichloro-2-butene
Concen: 669.04 ug/l
RT: 12.349 min Scan# 1727
Delta R.T. -0.036 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 75 Resp: 379655
Ion Ratio Lower Upper
75 100
53 0.0 75.0 112.6#
89 113.0 36.1 54.1#









#82
4-Chlorotoluene
Concen: 252.79 ug/l
RT: 12.825 min Scan# 1805
Delta R.T. -0.030 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 91 Resp: 1260497
Ion Ratio Lower Upper
91 100
126 3.9 16.9 50.7#



#83
tert-Butylbenzene
Concen: 97.33 ug/l
RT: 13.087 min Scan# 1848
Delta R.T. 0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 119 Resp: 494877
Ion Ratio Lower Upper
119 100
91 80.2 32.6 98.0
134 26.1 12.9 38.6









#84
1,2,4-Trimethylbenzene
Concen: 698.75 ug/l
RT: 13.264 min Scan# 1877
Delta R.T. 0.140 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:105 Resp: 4188852
Ion Ratio Lower Upper
105 100
120 0.0 22.4 67.3#



#85
sec-Butylbenzene
Concen: 590.60 ug/l
RT: 13.264 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion:105 Resp: 4188852
Ion Ratio Lower Upper
105 100
134 41.5 10.1 30.3#









#86
p-Isopropyltoluene
Concen: 286.23 ug/l
RT: 13.373 min Scan# 1895
Delta R.T. 0.000 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:119 Resp: 1846693
Ion Ratio Lower Upper
119 100
134 26.4 13.3 39.8
91 28.4 11.9 35.7



#89
n-Butylbenzene
Concen: 778.28 ug/l
RT: 13.703 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 91 Resp: 4792742
Ion Ratio Lower Upper
91 100
92 42.6 26.6 79.8
134 0.0 12.7 38.1#









#90
Hexachloroethane
Concen: 1005.83 ug/l
RT: 13.977 min Scan# 1994
Delta R.T. 0.012 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:117 Resp: 870639
Ion Ratio Lower Upper
117 100
201 0.0 36.6 109.9#



#92
1,2-Dibromo-3-Chloropropane
Concen: 40.29 ug/l
RT: 14.379 min Scan# 2060
Delta R.T. 0.019 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Tgt Ion: 75 Resp: 9996
Ion Ratio Lower Upper
75 100
155 14.6 40.6 121.8#
157 0.6 53.0 159.0#





#95
Naphthalene
Concen: 582.18 ug/l
RT: 15.239 min Scan# 2201
Delta R.T. 0.000 min
Lab File: VY004436.D
Acq: 15 Apr 2021 13:01

Instrument :
MSVOA_Y
ClientSampleId :

Tgt Ion:128 Resp: 2272849
Ion Ratio Lower Upper
128 100
127 13.1 10.2 15.2
129 11.3 8.8 13.2

