

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041521\
 Data File : VY004446.D
 Acq On : 15 Apr 2021 17:17
 Operator : SY/MD
 Sample : M2048-07
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-03-0-2.0

Manual Integrations
 APPROVED

MMDadoda
 4/16/2021 1:08:05 PM

Quant Time: Apr 16 05:05:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	220524	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	382883	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	365180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	170301	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	138947	56.71	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 113.42%			
35) Dibromofluoromethane	7.728	113	113314	53.33	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 106.66%			
50) Toluene-d8	10.185	98	470165	52.02	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 104.04%			
62) 4-Bromofluorobenzene	12.483	95	169463	55.25	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 110.50%			
Target Compounds						
11) Tert butyl alcohol	4.978	59	2889m	11.96	ug/l	Qvalue
20) Methylene Chloride	4.729	84	15192	6.42	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041521\
Data File : VY004446.D
Acq On : 15 Apr 2021 17:17
Operator : SY/MD
Sample : M2048-07
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-03-0-2.0

Quant Time: Apr 16 05:05:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
4/16/2021 1:08:05 PM

