

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041521\
 Data File : VY004455.D
 Acq On : 15 Apr 2021 20:47
 Operator : SY/MD
 Sample : M2049-08
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-09-5.5-6.0

Quant Time: Apr 16 05:07:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	192813	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	333681	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	319039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	147594	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	120736	56.36	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.72%
35) Dibromofluoromethane	7.728	113	98857	53.39	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.78%
50) Toluene-d8	10.185	98	413824	52.54	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.08%
62) 4-Bromofluorobenzene	12.483	95	144671	54.12	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.24%
Target Compounds						
51) 4-Methyl-2-Pentanone	10.075	43	2350	1.13	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041521\
Data File : VY004455.D
Acq On : 15 Apr 2021 20:47
Operator : SY/MD
Sample : M2049-08
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-09-5.5-6.0

Quant Time: Apr 16 05:07:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

