

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041621\
 Data File : VY004460.D
 Acq On : 16 Apr 2021 11:21
 Operator : SY/MD
 Sample : VY0416SBS01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0416SBS01

Quant Time: Apr 17 00:51:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	176727	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	286829	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	256835	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	119119	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	107612	54.80	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.60%	
35) Dibromofluoromethane	7.728	113	87117	54.73	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.46%	
50) Toluene-d8	10.185	98	368502	54.43	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.86%	
62) 4-Bromofluorobenzene	12.483	95	123213	53.63	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.26%	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.912	85	33865	20.68	ug/l 100
3) Chloromethane	2.119	50	48131	20.33	ug/l 96
4) Vinyl Chloride	2.259	62	54296	21.71	ug/l 99
5) Bromomethane	2.650	94	26653	21.85	ug/l 97
6) Chloroethane	2.802	64	31886	21.27	ug/l 100
7) Trichlorofluoromethane	3.131	101	65236	20.86	ug/l 99
8) Diethyl Ether	3.534	74	23105	21.86	ug/l 96
9) 1,1,2-Trichlorotrifluo...	3.912	101	37317	21.51	ug/l 100
10) Methyl Iodide	4.094	142	40600	19.56	ug/l 100
11) Tert butyl alcohol	4.966	59	23383	120.83	ug/l 100
12) 1,1-Dichloroethene	3.881	96	36543	21.30	ug/l 99
13) Acrolein	3.741	56	23013	106.98	ug/l 97
14) Allyl chloride	4.491	41	72231	22.02	ug/l 98
15) Acrylonitrile	5.174	53	65718	115.89	ug/l 99
16) Acetone	3.960	43	49706	108.37	ug/l 94
17) Carbon Disulfide	4.204	76	114346	21.65	ug/l 99
18) Methyl Acetate	4.491	43	30709	22.45	ug/l 99
19) Methyl tert-butyl Ether	5.235	73	113153	21.95	ug/l 98
20) Methylene Chloride	4.729	84	43655	23.02	ug/l 99
21) trans-1,2-Dichloroethene	5.235	96	41195	21.22	ug/l 95
22) Diisopropyl ether	6.131	45	152697	22.02	ug/l 99
23) Vinyl Acetate	6.070	43	533417	112.99	ug/l 100
24) 1,1-Dichloroethane	6.027	63	76640	21.63	ug/l 99
25) 2-Butanone	6.996	43	87396	116.85	ug/l 95
26) 2,2-Dichloropropane	6.990	77	68106	21.42	ug/l 100
27) cis-1,2-Dichloroethene	6.996	96	46506	21.41	ug/l 98
28) Bromochloromethane	7.344	49	33067	20.54	ug/l 98
29) Tetrahydrofuran	7.356	42	61391	114.88	ug/l 100
30) Chloroform	7.515	83	74503	21.40	ug/l 98
31) Cyclohexane	7.789	56	76764	21.33	ug/l 96
32) 1,1,1-Trichloroethane	7.710	97	66364	21.01	ug/l 100
36) 1,1-Dichloropropene	7.923	75	60016	21.06	ug/l 100
37) Ethyl Acetate	7.082	43	43126	23.50	ug/l 98
38) Carbon Tetrachloride	7.911	117	52569	21.34	ug/l 97
39) Methylcyclohexane	9.185	83	74757	21.30	ug/l 99
40) Benzene	8.167	78	172756	21.57	ug/l 99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041621\
 Data File : VY004460.D
 Acq On : 16 Apr 2021 11:21
 Operator : SY/MD
 Sample : VY0416SBS01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0416SBS01

Quant Time: Apr 17 00:51:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.350	41	57244	51.46	ug/l #	100
42) 1,2-Dichloroethane	8.246	62	52883	21.05	ug/l	99
43) Isopropyl Acetate	8.277	43	78674	22.61	ug/l	99
44) Trichloroethene	8.941	130	46694	21.38	ug/l	99
45) 1,2-Dichloropropane	9.222	63	45757	22.13	ug/l	100
46) Dibromomethane	9.313	93	24125	21.44	ug/l	98
47) Bromodichloromethane	9.502	83	58244	21.38	ug/l	99
48) Methyl methacrylate	9.301	41	35494	22.15	ug/l	98
49) 1,4-Dioxane	9.307	88	5883	400.24	ug/l	98
51) 4-Methyl-2-Pentanone	10.075	43	208824	116.84	ug/l	98
52) Toluene	10.246	92	108649	20.96	ug/l	99
53) t-1,3-Dichloropropene	10.471	75	64850	21.58	ug/l	100
54) cis-1,3-Dichloropropene	9.935	75	72859	21.35	ug/l	99
55) 1,1,2-Trichloroethane	10.648	97	33618	21.68	ug/l	95
56) Ethyl methacrylate	10.514	69	47916	21.31	ug/l	97
57) 1,3-Dichloropropane	10.794	76	60875	22.00	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.789	63	123180	107.05	ug/l	99
59) 2-Hexanone	10.837	43	139101	116.10	ug/l	99
60) Dibromochloromethane	10.990	129	38903	21.11	ug/l	98
61) 1,2-Dibromoethane	11.093	107	32887	22.00	ug/l	100
64) Tetrachloroethene	10.721	164	47846	21.40	ug/l	98
65) Chlorobenzene	11.520	112	111032	21.18	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.593	131	39792	21.27	ug/l	99
67) Ethyl Benzene	11.593	91	207923	21.20	ug/l	97
68) m/p-Xylenes	11.709	106	154965	42.53	ug/l	99
69) o-Xylene	12.032	106	72110	21.12	ug/l	99
70) Styrene	12.050	104	118970	21.03	ug/l	99
71) Bromoform	12.215	173	23531	21.21	ug/l #	99
73) Isopropylbenzene	12.337	105	195911	21.28	ug/l	100
74) N-amyl acetate	12.148	43	68949	22.25	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.587	83	39385	22.62	ug/l	99
76) 1,2,3-Trichloropropane	12.642	75	54540	42.66	ug/l #	100
77) Bromobenzene	12.611	156	43779	21.22	ug/l	100
78) n-propylbenzene	12.672	91	235374	21.65	ug/l	99
79) 2-Chlorotoluene	12.758	91	127768	21.48	ug/l	99
80) 1,3,5-Trimethylbenzene	12.819	105	164045	21.93	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.386	75	15708	21.98	ug/l	97
82) 4-Chlorotoluene	12.861	91	134044	21.35	ug/l	100
83) tert-Butylbenzene	13.081	119	136178	21.27	ug/l	99
84) 1,2,4-Trimethylbenzene	13.123	105	162608	21.54	ug/l	100
85) sec-Butylbenzene	13.257	105	191815	21.47	ug/l	99
86) p-Isopropyltoluene	13.373	119	171892	21.15	ug/l	99
87) 1,3-Dichlorobenzene	13.373	146	82248	21.34	ug/l	99
88) 1,4-Dichlorobenzene	13.453	146	82276	21.21	ug/l	99
89) n-Butylbenzene	13.703	91	165413	21.33	ug/l	98
90) Hexachloroethane	13.965	117	24062	22.07	ug/l	96
91) 1,2-Dichlorobenzene	13.745	146	72085	21.07	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.367	75	6242	19.97	ug/l	95
93) 1,2,4-Trichlorobenzene	15.013	180	46672	20.64	ug/l	99
94) Hexachlorobutadiene	15.117	225	24071	20.39	ug/l	98
95) Naphthalene	15.239	128	101264	20.60	ug/l	100
96) 1,2,3-Trichlorobenzene	15.428	180	40503	20.30	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041621\
Data File : VY004460.D
Acq On : 16 Apr 2021 11:21
Operator : SY/MD
Sample : VY0416SBS01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Quant Time: Apr 17 00:51:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 17 00:51:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.801 min Scan# 981
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:168 Resp: 176727
Ion Ratio Lower Upper
168 100
99 53.1 45.5 68.3



#2
Dichlorodifluoromethane
Concen: 20.68 ug/l
RT: 1.912 min Scan# 15
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 85 Resp: 33865
Ion Ratio Lower Upper
85 100
87 33.8 16.8 50.4









#3
Chloromethane
Concen: 20.33 ug/l
RT: 2.119 min Scan# 49
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 50 Resp: 48131
Ion Ratio Lower Upper
50 100
52 30.7 26.6 39.8



#4
Vinyl Chloride
Concen: 21.71 ug/l
RT: 2.259 min Scan# 72
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 62 Resp: 54296
Ion Ratio Lower Upper
62 100
64 32.1 25.1 37.7









#5
Bromomethane
Concen: 21.85 ug/l
RT: 2.650 min Scan# 136
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01

Tgt Ion: 94 Resp: 26653
Ion Ratio Lower Upper
94 100
96 94.0 72.6 108.8



#6
Chloroethane
Concen: 21.27 ug/l
RT: 2.802 min Scan# 161
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 64 Resp: 31886
Ion Ratio Lower Upper
64 100
66 33.0 26.3 39.5









#7
Trichlorofluoromethane
Concen: 20.86 ug/l
RT: 3.131 min Scan# 215
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:101 Resp: 65236
Ion Ratio Lower Upper
101 100
103 63.5 51.2 76.8



#8
Diethyl Ether
Concen: 21.86 ug/l
RT: 3.534 min Scan# 281
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 74 Resp: 23105
Ion Ratio Lower Upper
74 100
45 108.1 52.3 156.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.51 ug/l
RT: 3.912 min Scan# 343
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:101 Resp: 37317
Ion Ratio Lower Upper
101 100
85 46.0 36.9 55.3
151 77.0 62.0 93.0



#10
Methyl Iodide
Concen: 19.56 ug/l
RT: 4.094 min Scan# 373
Delta R.T. -0.012 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21


Tgt Ion:142 Resp: 40600
Ion Ratio Lower Upper
142 100
127 45.6 36.2 54.4
141 14.3 11.6 17.4








#11
Tert butyl alcohol
Concen: 120.83 ug/l
RT: 4.966 min Scan# 516
Delta R.T. -0.012 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01







Tgt Ion: 59 Resp: 23383
Ion Ratio Lower Upper
59 100
57 10.0 8.1 12.1



#12
1,1-Dichloroethene
Concen: 21.30 ug/l
RT: 3.881 min Scan# 338
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21


Tgt Ion: 96 Resp: 36543
Ion Ratio Lower Upper
96 100
61 166.3 132.1 198.1
98 61.9 48.2 72.2








#13
Acrolein
Concen: 106.98 ug/l
RT: 3.741 min Scan# 315
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01







Tgt Ion: 56 Resp: 23013
Ion Ratio Lower Upper
56 100
55 71.6 55.4 83.2



#14
Allyl chloride
Concen: 22.02 ug/l
RT: 4.491 min Scan# 438
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 41 Resp: 72231
Ion Ratio Lower Upper
41 100
39 68.9 56.9 85.3
76 32.6 26.8 40.2









#15
Acrylonitrile
Concen: 115.89 ug/l
RT: 5.174 min Scan# 550
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 53	Resp: 65718
Ion Ratio	Lower Upper
53 100	
52 82.2	64.9 97.3
51 37.3	29.3 43.9



#16
Acetone
Concen: 108.37 ug/l
RT: 3.960 min Scan# 351
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 43	Resp: 49706
Ion Ratio	Lower Upper
43 100	
58 27.8	24.6 37.0









#17
Carbon Disulfide
Concen: 21.65 ug/l
RT: 4.204 min Scan# 391
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 76 Resp: 114346
Ion Ratio Lower Upper
76 100
78 8.8 7.4 11.2



#18
Methyl Acetate
Concen: 22.45 ug/l
RT: 4.491 min Scan# 438
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 43 Resp: 30709
Ion Ratio Lower Upper
43 100
74 21.0 17.2 25.8





#19

Methyl tert-butyl Ether

Concen: 21.95 ug/l

RT: 5.235 min Scan# 560

Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21

Tgt Ion: 73 Resp: 113153

Ion Ratio Lower Upper

73 100





57 23.9 18.4 27.6

Instrument :

MSVOA_Y

ClientSampleId :

VY0416SBS01



#20

Methylene Chloride

Concen: 23.02 ug/l

RT: 4.729 min Scan# 477

Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21

Tgt Ion: 84 Resp: 43655




Ion Ratio Lower Upper


84 100

49 139.6 111.2 166.8

51 39.7 33.4 50.2


86 64.5 52.7 79.1








#21
trans-1,2-Dichloroethene
Concen: 21.22 ug/l
RT: 5.235 min Scan# 560
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01







Tgt Ion: 96 Resp: 41195
Ion Ratio Lower Upper
96 100
61 138.2 116.5 174.7
98 61.9 52.0 78.0



#22
Diisopropyl ether
Concen: 22.02 ug/l
RT: 6.131 min Scan# 707
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 45 Resp: 152697
Ion Ratio Lower Upper
45 100
43 58.4 46.3 69.5
87 24.7 20.3 30.5
59 10.5 8.6 12.8









#23
 Vinyl Acetate
 Concen: 112.99 ug/l
 RT: 6.070 min Scan# 697
 Delta R.T. -0.006 min
 Lab File: VY004460.D
 Acq: 16 Apr 2021 11:21




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0416SBS01


Tgt Ion: 43 Resp: 533417
 Ion Ratio Lower Upper
 43 100
 86 9.4 7.6 11.4



#24
 1,1-Dichloroethane
 Concen: 21.63 ug/l
 RT: 6.027 min Scan# 690
 Delta R.T. -0.006 min
 Lab File: VY004460.D
 Acq: 16 Apr 2021 11:21

Tgt Ion: 63 Resp: 76640
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.3 9.9
 100 4.6 2.0 6.0









#25
2-Butanone
Concen: 116.85 ug/l
RT: 6.996 min Scan# 849
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 43 Resp: 87396
Ion Ratio Lower Upper
43 100
72 22.8 20.2 30.2



#26
2,2-Dichloropropane
Concen: 21.42 ug/l
RT: 6.990 min Scan# 848
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 77 Resp: 68106
Ion Ratio Lower Upper
77 100
97 22.2 11.1 33.3





#27

cis-1,2-Dichloroethene

Concen: 21.41 ug/l

RT: 6.996 min Scan# 849

Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21





Instrument :

MSVOA_Y

ClientSampleId :

VY0416SBS01

Tgt Ion:	96	Resp:	46506
Ion	Ratio	Lower	Upper
96	100		
61	154.5	0.0	303.2
98	63.8	0.0	130.0



#28

Bromochloromethane

Concen: 20.54 ug/l




RT: 7.344 min Scan# 906


Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21

Tgt Ion:	49	Resp:	33067
Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.8
130	65.5	53.9	80.9





#29

Tetrahydrofuran

Concen: 114.88 ug/l

RT: 7.356 min Scan# 908





Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21

Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01

Tgt Ion:	42	Resp:	61391
Ion	Ratio	Lower	Upper
42	100		
72	38.5	31.0	46.4
71	35.9	28.8	43.2



#30

Chloroform

Concen: 21.40 ug/l




RT: 7.515 min Scan# 934


Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21


Tgt Ion:	83	Resp:	74503
Ion	Ratio	Lower	Upper
83	100		
85	66.0	51.3	76.9








#31
Cyclohexane
Concen: 21.33 ug/l
RT: 7.789 min Scan# 979
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21


Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01





Tgt Ion: 56 Resp: 76764
Ion Ratio Lower Upper
56 100
69 36.1 25.8 38.6
84 82.8 68.2 102.4




#32
1,1,1-Trichloroethane
Concen: 21.01 ug/l
RT: 7.710 min Scan# 966
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21



Tgt Ion: 97 Resp: 66364
Ion Ratio Lower Upper
97 100
99 63.7 51.2 76.8
61 45.7 36.2 54.2









#33
 1,2-Dichloroethane-d4
 Concen: 54.80 ug/l
 RT: 8.149 min Scan# 1038
 Delta R.T. -0.006 min
 Lab File: VY004460.D
 Acq: 16 Apr 2021 11:21




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0416SBS01


Tgt Ion: 65 Resp: 107612
 Ion Ratio Lower Upper
 65 100
 67 53.4 0.0 106.4



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.697 min Scan# 1128
 Delta R.T. 0.000 min
 Lab File: VY004460.D
 Acq: 16 Apr 2021 11:21

Tgt Ion:114 Resp: 286829
 Ion Ratio Lower Upper
 114 100
 63 20.8 0.0 42.4
 88 15.6 0.0 30.8









#35
Dibromofluoromethane
Concen: 54.73 ug/l
RT: 7.728 min Scan# 969
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:	113	Resp:	87117
Ion Ratio	Lower	Upper	
113	100		
111	102.0	82.6	123.8
192	19.4	16.1	24.1



#36
1,1-Dichloropropene
Concen: 21.06 ug/l
RT: 7.923 min Scan# 1001
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:	75	Resp:	60016
Ion Ratio	Lower	Upper	
75	100		
110	34.7	17.3	51.7
77	30.8	24.3	36.5





#37

Ethyl Acetate

Concen: 23.50 ug/l

RT: 7.082 min Scan# 863

Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21





Instrument :

MSVOA_Y

ClientSampleId :

VY0416SBS01

Tgt Ion: 43	Resp: 43126
Ion Ratio	Lower Upper
43 100	
61 13.9	10.4 15.6
70 9.0	7.5 11.3



#38

Carbon Tetrachloride

Concen: 21.34 ug/l




RT: 7.911 min Scan# 999


Delta R.T. 0.000 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21

Tgt Ion: 117	Resp: 52569
Ion Ratio	Lower Upper
117 100	
119 90.8	75.8 113.8
121 31.6	25.3 37.9









#39
Methylcyclohexane
Concen: 21.30 ug/l
RT: 9.185 min Scan# 1208
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 83 Resp: 74757
Ion Ratio Lower Upper
83 100
55 79.5 63.0 94.4
98 49.7 39.2 58.8



#40
Benzene
Concen: 21.57 ug/l
RT: 8.167 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 78 Resp: 172756
Ion Ratio Lower Upper
78 100
77 23.6 19.4 29.0









#41
Methacrylonitrile
Concen: 51.46 ug/l
RT: 7.350 min Scan# 907
Delta R.T. 0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 41	Resp: 57244
Ion Ratio	Lower Upper
41 100	
39 0.0	0.0 0.0
67 0.0	0.0 0.0
52 0.0	0.0 0.0



#42
1,2-Dichloroethane
Concen: 21.05 ug/l
RT: 8.246 min Scan# 1054
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 62	Resp: 52883
Ion Ratio	Lower Upper
62 100	
98 9.6	0.0 18.2





#43

Isopropyl Acetate

Concen: 22.61 ug/l

RT: 8.277 min Scan# 1059

Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21





Instrument :

MSVOA_Y

ClientSampleId :

VY0416SBS01

Tgt Ion: 43	Resp: 78674
Ion Ratio	Lower Upper
43 100	
61 25.9	20.9 31.3
87 11.9	9.1 13.7



#44

Trichloroethene

Concen: 21.38 ug/l




RT: 8.941 min Scan# 1168


Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21

Tgt Ion: 130	Resp: 46694
Ion Ratio	Lower Upper
130 100	
95 94.3	0.0 190.4





#45

1,2-Dichloropropane

Concen: 22.13 ug/l

RT: 9.222 min Scan# 1214

Delta R.T. -0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21

Tgt Ion: 63 Resp: 45757

Ion Ratio Lower Upper

63 100





65 30.7 24.4 36.6

Instrument :

MSVOA_Y

ClientSampleId :

VY0416SBS01



#46

Dibromomethane

Concen: 21.44 ug/l

RT: 9.313 min Scan# 1229

Delta R.T. 0.000 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21




Tgt Ion: 93 Resp: 24125


Ion Ratio Lower Upper

93 100

95 83.9 65.8 98.8

174 94.6 77.8 116.6









#47
Bromodichloromethane
Concen: 21.38 ug/l
RT: 9.502 min Scan# 1260
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 83 Resp: 58244
Ion Ratio Lower Upper
83 100
85 63.9 50.7 76.1
127 7.2 6.6 10.0



#48
Methyl methacrylate
Concen: 22.15 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 41 Resp: 35494
Ion Ratio Lower Upper
41 100
69 76.3 62.6 94.0
39 59.5 47.2 70.8









#49
1,4-Dioxane
Concen: 400.24 ug/l
RT: 9.307 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 88 Resp: 5883
Ion Ratio Lower Upper
88 100
43 41.4 31.7 47.5
58 68.9 55.7 83.5



#50
Toluene-d8
Concen: 54.43 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 98 Resp: 368502
Ion Ratio Lower Upper
98 100
100 64.7 51.8 77.6









#51
4-Methyl-2-Pentanone
Concen: 116.84 ug/l
RT: 10.075 min Scan# 1354
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 43 Resp: 208824
Ion Ratio Lower Upper
43 100
58 35.5 29.2 43.8



#52
Toluene
Concen: 20.96 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 92 Resp: 108649
Ion Ratio Lower Upper
92 100
91 173.2 137.8 206.6









#53
t-1,3-Dichloropropene
Concen: 21.58 ug/l
RT: 10.471 min Scan# 1419
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 75 Resp: 64850
Ion Ratio Lower Upper
75 100
77 31.3 25.2 37.8



#54
cis-1,3-Dichloropropene
Concen: 21.35 ug/l
RT: 9.935 min Scan# 1331
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 75 Resp: 72859
Ion Ratio Lower Upper
75 100
77 31.8 24.6 37.0
39 52.4 42.1 63.1









#55
1,1,2-Trichloroethane
Concen: 21.68 ug/l
RT: 10.648 min Scan# 1448
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:	97	Resp:	33618
Ion	Ratio	Lower	Upper
97	100		
83	93.0	68.4	102.6
85	54.7	46.2	69.4
99	59.5	48.8	73.2



#56
Ethyl methacrylate
Concen: 21.31 ug/l
RT: 10.514 min Scan# 1426
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:	69	Resp:	47916
Ion	Ratio	Lower	Upper
69	100		
41	77.0	59.2	88.8
39	39.4	31.0	46.6









#57
1,3-Dichloropropane
Concen: 22.00 ug/l
RT: 10.794 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 76 Resp: 60875
Ion Ratio Lower Upper
76 100
78 32.0 25.9 38.9



#58
2-Chloroethyl Vinyl ether
Concen: 107.05 ug/l
RT: 9.789 min Scan# 1307
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 63 Resp: 123180
Ion Ratio Lower Upper
63 100
106 26.8 21.7 32.5









#59
2-Hexanone
Concen: 116.10 ug/l
RT: 10.837 min Scan# 1479
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 43 Resp: 139101
Ion Ratio Lower Upper
43 100
58 50.2 25.4 76.4



#60
Dibromochloromethane
Concen: 21.11 ug/l
RT: 10.990 min Scan# 1504
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 129 Resp: 38903
Ion Ratio Lower Upper
129 100
127 79.3 38.9 116.7









#61
1,2-Dibromoethane
Concen: 22.00 ug/l
RT: 11.093 min Scan# 1521
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01

Tgt Ion:107 Resp: 32887
Ion Ratio Lower Upper
107 100
109 93.9 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 53.63 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 95 Resp: 123213
Ion Ratio Lower Upper
95 100
174 79.9 0.0 162.4
176 78.5 0.0 156.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.496 min Scan# 1587
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:117 Resp: 256835
Ion Ratio Lower Upper
117 100
82 55.5 44.9 67.3
119 31.5 25.4 38.0



#64
Tetrachloroethene
Concen: 21.40 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:164 Resp: 47846
Ion Ratio Lower Upper
164 100
166 123.3 102.1 153.1
129 92.9 74.8 112.2
131 89.2 72.2 108.4









#65
Chlorobenzene
Concen: 21.18 ug/l
RT: 11.520 min Scan# 1591
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01

Tgt Ion:112 Resp: 111032
Ion Ratio Lower Upper
112 100
114 33.6 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 21.27 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:131 Resp: 39792
Ion Ratio Lower Upper
131 100
133 94.5 47.6 142.9
119 66.4 32.6 98.0









#67
Ethyl Benzene
Concen: 21.20 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. -0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 91 Resp: 207923
Ion Ratio Lower Upper
91 100
106 29.9 25.1 37.7



#68
m/p-Xylenes
Concen: 42.53 ug/l
RT: 11.709 min Scan# 1622
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:106 Resp: 154965
Ion Ratio Lower Upper
106 100
91 202.7 163.2 244.8









#69
o-Xylene
Concen: 21.12 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:106 Resp: 72110
Ion Ratio Lower Upper
106 100
91 213.1 107.4 322.2



#70
Styrene
Concen: 21.03 ug/l
RT: 12.050 min Scan# 1678
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:104 Resp: 118970
Ion Ratio Lower Upper
104 100
78 49.7 39.9 59.9
103 55.7 43.4 65.2









#71
Bromoform
Concen: 21.21 ug/l
RT: 12.215 min Scan# 1705
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:173 Resp: 23531
Ion Ratio Lower Upper
173 100
175 47.9 24.3 72.8
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:152 Resp: 119119
Ion Ratio Lower Upper
152 100
115 59.6 29.6 88.8
150 164.7 0.0 348.4









#73
Isopropylbenzene
Concen: 21.28 ug/l
RT: 12.337 min Scan# 1725
Delta R.T. 0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:105 Resp: 195911
Ion Ratio Lower Upper
105 100
120 26.1 13.2 39.5



#74
N-ethyl acetate
Concen: 22.25 ug/l
RT: 12.148 min Scan# 1694
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 43 Resp: 68949
Ion Ratio Lower Upper
43 100
70 37.8 30.4 45.6
55 21.7 17.4 26.0
61 22.3 18.2 27.2









#75
1,1,2,2-Tetrachloroethane
Concen: 22.62 ug/l
RT: 12.587 min Scan# 1766
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:	83	Resp:	39385
Ion Ratio	Lower	Upper	
83	100		
131	9.9	5.1	15.4
85	63.6	31.6	94.8



#76
1,2,3-Trichloropropane
Concen: 42.66 ug/l
RT: 12.642 min Scan# 1775
Delta R.T. 0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:	75	Resp:	54540
Ion Ratio	Lower	Upper	
75	100		
77	0.0	0.0	0.0









#77
Bromobenzene
Concen: 21.22 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:156 Resp: 43779
Ion Ratio Lower Upper
156 100
77 196.3 98.6 295.7
158 97.1 48.4 145.2



#78
n-propylbenzene
Concen: 21.65 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 91 Resp: 235374
Ion Ratio Lower Upper
91 100
120 22.5 11.5 34.4









#79
2-Chlorotoluene
Concen: 21.48 ug/l
RT: 12.758 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 91 Resp: 127768
Ion Ratio Lower Upper
91 100
126 33.7 16.7 50.1



#80
1,3,5-Trimethylbenzene
Concen: 21.93 ug/l
RT: 12.819 min Scan# 1804
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 105 Resp: 164045
Ion Ratio Lower Upper
105 100
120 48.2 24.6 74.0









#81
trans-1,4-Dichloro-2-butene
Concen: 21.98 ug/l
RT: 12.386 min Scan# 1733
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 75 Resp: 15708
Ion Ratio Lower Upper
75 100
53 93.2 75.0 112.6
89 40.1 36.1 54.1



#82
4-Chlorotoluene
Concen: 21.35 ug/l
RT: 12.861 min Scan# 1811
Delta R.T. 0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 91 Resp: 134044
Ion Ratio Lower Upper
91 100
126 34.0 16.9 50.7









#83
tert-Butylbenzene
Concen: 21.27 ug/l
RT: 13.081 min Scan# 1847
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:119 Resp: 136178
Ion Ratio Lower Upper
119 100
91 65.2 32.6 98.0
134 26.5 12.9 38.6



#84
1,2,4-Trimethylbenzene
Concen: 21.54 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:105 Resp: 162608
Ion Ratio Lower Upper
105 100
120 44.8 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 21.47 ug/l
 RT: 13.257 min Scan# 1876
 Delta R.T. 0.000 min
 Lab File: VY004460.D
 Acq: 16 Apr 2021 11:21




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0416SBS01


Tgt Ion:105 Resp: 191815
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.1 30.3



#86
 p-Isopropyltoluene
 Concen: 21.15 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.000 min
 Lab File: VY004460.D
 Acq: 16 Apr 2021 11:21

Tgt Ion:119 Resp: 171892
 Ion Ratio Lower Upper
 119 100
 134 26.3 13.3 39.8
 91 24.1 11.9 35.7





#87

1,3-Dichlorobenzene

Concen: 21.34 ug/l

RT: 13.373 min Scan# 1895

Delta R.T. 0.006 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21





Tgt Ion:146 Resp: 82248

Ion Ratio Lower Upper

146 100

111 40.4 20.4 61.2

148 62.8 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 21.21 ug/l

RT: 13.453 min Scan# 1908

Delta R.T. 0.000 min

Lab File: VY004460.D

Acq: 16 Apr 2021 11:21




Tgt Ion:146 Resp: 82276


Ion Ratio Lower Upper

146 100

111 39.8 19.6 58.8

148 64.8 31.9 95.5









#89
n-Butylbenzene
Concen: 21.33 ug/l
RT: 13.703 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion: 91 Resp: 165413
Ion Ratio Lower Upper
91 100
92 54.4 26.6 79.8
134 26.0 12.7 38.1



#90
Hexachloroethane
Concen: 22.07 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 117 Resp: 24062
Ion Ratio Lower Upper
117 100
201 69.8 36.6 109.9









#91
1,2-Dichlorobenzene
Concen: 21.07 ug/l
RT: 13.745 min Scan# 1956
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01


Tgt Ion:146 Resp: 72085
Ion Ratio Lower Upper
146 100
111 41.9 20.9 62.7
148 64.4 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.97 ug/l
RT: 14.367 min Scan# 2058
Delta R.T. 0.006 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion: 75 Resp: 6242
Ion Ratio Lower Upper
75 100
155 86.7 40.6 121.8
157 110.7 53.0 159.0









#93
1,2,4-Trichlorobenzene
Concen: 20.64 ug/l
RT: 15.013 min Scan# 2164
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21




Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01

Tgt Ion:180 Resp: 46672
Ion Ratio Lower Upper
180 100
182 93.2 47.3 141.8
145 30.9 15.5 46.5



#94
Hexachlorobutadiene
Concen: 20.39 ug/l
RT: 15.117 min Scan# 2181
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:225 Resp: 24071
Ion Ratio Lower Upper
225 100
223 62.3 31.3 93.9
227 61.5 32.2 96.6









#95
Naphthalene
Concen: 20.60 ug/l
RT: 15.239 min Scan# 2201
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Instrument :
MSVOA_Y
ClientSampleId :
VY0416SBS01

Tgt Ion:128 Resp: 101264
Ion Ratio Lower Upper
128 100
127 12.8 10.2 15.2
129 11.0 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 20.30 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.000 min
Lab File: VY004460.D
Acq: 16 Apr 2021 11:21

Tgt Ion:180 Resp: 40503
Ion Ratio Lower Upper
180 100
182 93.1 47.6 142.8
145 32.3 16.3 48.8

