

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041621\
 Data File : VY004480.D
 Acq On : 16 Apr 2021 19:15
 Operator : SY/MD
 Sample : M2048-10
 Misc : 5.08G/5ML/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-04-0-2.0

Quant Time: Apr 17 00:57:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	149379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	260924	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	248155	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	112253	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.154	65	95082	57.29	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.58%
35) Dibromofluoromethane	7.728	113	79631	54.99	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.98%
50) Toluene-d8	10.184	98	324925	52.75	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.50%
62) 4-Bromofluorobenzene	12.483	95	110068	52.66	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.32%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.972	59	2220	13.57	ug/l #	87
16) Acetone	3.966	43	10531	11.55	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041621\
Data File : VY004480.D
Acq On : 16 Apr 2021 19:15
Operator : SY/MD
Sample : M2048-10
Misc : 5.08G/5ML/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-04-0-2.0

Quant Time: Apr 17 00:57:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

