

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041621\
 Data File : VY004484.D
 Acq On : 16 Apr 2021 20:47
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 17 00:58:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	69	0.00
2 T	Dichlorodifluoromethane	0.463	0.458	1.1	65	0.00
3 P	Chloromethane	0.670	0.682	-1.8	71	0.00
4 C	Vinyl Chloride	0.707	0.789	-11.6#	77	0.00
5 T	Bromomethane	0.345	0.376	-9.0	76	0.00
6 T	Chloroethane	0.424	0.464	-9.4	73	0.00
7 T	Trichlorofluoromethane	0.885	0.891	-0.7	68	0.00
8 T	Diethyl Ether	0.299	0.348	-16.4	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.491	0.499	-1.6	68	0.00
10 T	Methyl Iodide	0.587	0.589	-0.3	68	0.00
11 T	Tert butyl alcohol	0.055	0.068	-23.6	96	0.00
12 CM	1,1-Dichloroethene	0.485	0.497	-2.5#	69	0.00
13 T	Acrolein	0.061	0.079	-29.5#	95	0.00
14 T	Allyl chloride	0.928	1.015	-9.4	76	0.00
15 T	Acrylonitrile	0.160	0.198	-23.8	89	0.00
16 T	Acetone	0.133	0.161	-21.1	96	0.00
17 T	Carbon Disulfide	1.494	1.525	-2.1	68	0.00
18 T	Methyl Acetate	0.387	0.492	-27.1#	96	0.00
19 T	Methyl tert-butyl Ether	1.458	1.691	-16.0	81	0.00
20 T	Methylene Chloride	0.537	0.647	-20.5	84	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.593	-8.0	73	0.00
22 T	Diisopropyl ether	1.962	2.280	-16.2	79	0.00
23 T	Vinyl Acetate	1.336	1.600	-19.8	83	0.00
24 P	1,1-Dichloroethane	1.002	1.116	-11.4	75	0.00
25 T	2-Butanone	0.212	0.279	-31.6#	95	0.00
26 T	2,2-Dichloropropane	0.899	0.866	3.7	65	0.00
27 T	cis-1,2-Dichloroethene	0.614	0.682	-11.1	75	0.00
28 T	Bromochloromethane	0.456	0.578	-26.8#	82	0.00
29 T	Tetrahydrofuran	0.151	0.193	-27.8#	94	0.00
30 C	Chloroform	0.985	1.097	-11.4#	76	0.00
31 T	Cyclohexane	1.018	0.990	2.8	68	0.00
32 T	1,1,1-Trichloroethane	0.894	0.935	-4.6	70	0.00
33 S	1,2-Dichloroethane-d4	0.556	0.654	-17.6	78	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	73	0.00
35 S	Dibromofluoromethane	0.277	0.298	-7.6	74	0.00
36 T	1,1-Dichloropropene	0.497	0.491	1.2	70	0.00
37 T	Ethyl Acetate	0.320	0.376	-17.5	92	0.00
38 T	Carbon Tetrachloride	0.430	0.441	-2.6	70	0.00
39 T	Methylcyclohexane	0.612	0.584	4.6	67	0.00
40 TM	Benzene	1.396	1.472	-5.4	76	0.00
41 T	Methacrylonitrile	0.194	0.189	2.6	67	-0.02
42 TM	1,2-Dichloroethane	0.438	0.466	-6.4	79	0.00
43 T	Isopropyl Acetate	0.607	0.718	-18.3	89	0.00
44 TM	Trichloroethene	0.381	0.387	-1.6	72	0.00
45 C	1,2-Dichloropropane	0.360	0.393	-9.2#	78	0.00
46 T	Dibromomethane	0.196	0.221	-12.8	83	0.00
47 T	Bromodichloromethane	0.475	0.519	-9.3	78	0.00
48 T	Methyl methacrylate	0.279	0.332	-19.0	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041621\
 Data File : VY004484.D
 Acq On : 16 Apr 2021 20:47
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 17 00:58:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	107	0.00
50 S	Toluene-d8	1.180	1.278	-8.3	74	0.00
51 T	4-Methyl-2-Pentanone	0.312	0.384	-23.1	94	0.00
52 CM	Toluene	0.904	0.921	-1.9#	73	0.00
53 T	t-1,3-Dichloropropene	0.524	0.560	-6.9	78	0.00
54 T	cis-1,3-Dichloropropene	0.595	0.623	-4.7	76	0.00
55 T	1,1,2-Trichloroethane	0.270	0.301	-11.5	83	0.00
56 T	Ethyl methacrylate	0.392	0.451	-15.1	84	0.00
57 T	1,3-Dichloropropane	0.482	0.544	-12.9	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.239	-18.9	83	0.00
59 T	2-Hexanone	0.209	0.266	-27.3#	96	0.00
60 T	Dibromochloromethane	0.321	0.358	-11.5	80	0.00
61 T	1,2-Dibromoethane	0.261	0.291	-11.5	82	0.00
62 S	4-Bromofluorobenzene	0.401	0.432	-7.7	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.435	0.427	1.8	73	0.00
65 PM	Chlorobenzene	1.020	1.037	-1.7	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.380	-4.4	77	0.00
67 C	Ethyl Benzene	1.909	1.903	0.3#	74	0.00
68 T	m/p-Xylenes	0.709	0.713	-0.6	74	0.00
69 T	o-Xylene	0.665	0.674	-1.4	75	0.00
70 T	Styrene	1.101	1.156	-5.0	76	0.00
71 P	Bromoform	0.216	0.235	-8.8	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
73 T	Isopropylbenzene	3.864	3.787	2.0	72	0.00
74 T	N-amyl acetate	1.301	1.508	-15.9	90	0.00
75 P	1,1,2,2-Tetrachloroethane	0.731	0.814	-11.4	88	0.00
76 T	1,2,3-Trichloropropane	0.537	0.592	-10.2	84	0.00
77 T	Bromobenzene	0.866	0.876	-1.2	78	0.00
78 T	n-propylbenzene	4.564	4.480	1.8	72	0.00
79 T	2-Chlorotoluene	2.496	2.469	1.1	74	0.00
80 T	1,3,5-Trimethylbenzene	3.140	3.134	0.2	74	0.00
81 T	trans-1,4-Dichloro-2-butene	0.300	0.316	-5.3	84	0.00
82 T	4-Chlorotoluene	2.636	2.635	0.0	75	0.00
83 T	tert-Butylbenzene	2.688	2.684	0.1	74	0.00
84 T	1,2,4-Trimethylbenzene	3.169	3.218	-1.5	76	0.00
85 T	sec-Butylbenzene	3.749	3.708	1.1	72	0.00
86 T	p-Isopropyltoluene	3.411	3.343	2.0	72	0.00
87 T	1,3-Dichlorobenzene	1.618	1.642	-1.5	77	0.00
88 T	1,4-Dichlorobenzene	1.628	1.620	0.5	76	0.00
89 T	n-Butylbenzene	3.255	3.149	3.3	71	0.00
90 T	Hexachloroethane	0.458	0.485	-5.9	78	0.00
91 T	1,2-Dichlorobenzene	1.436	1.478	-2.9	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.131	0.145	-10.7	93	0.00
93 T	1,2,4-Trichlorobenzene	0.949	0.923	2.7	76	0.00
94 T	Hexachlorobutadiene	0.496	0.473	4.6	70	0.00
95 T	Naphthalene	2.064	2.196	-6.4	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041621\
Data File : VY004484.D
Acq On : 16 Apr 2021 20:47
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 17 00:58:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.838	0.836	0.2	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6