



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041621\
Data File : VY004457.D
Acq On : 16 Apr 2021 09:25
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Title : SW846 8260
Last Update : Tue Apr 13 03:38:02 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.0	18659	PASS
75	95	30	60	56.3	47677	PASS
95	95	100	100	100.0	84680	PASS
96	95	5	9	7.3	6152	PASS
173	174	0.00	2	1.1	772	PASS
174	95	50	100	81.1	68635	PASS
175	174	5	9	7.7	5301	PASS
176	174	95	101	96.5	66216	PASS
177	176	5	9	6.8	4520	PASS