



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041624\
Data File : VY017939.D
Acq On : 16 Apr 2024 10:15
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Title : SW846 8260
Last Update : Wed Apr 17 06:24:42 2024



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.7	20616	PASS
75	95	30	60	56.3	53520	PASS
95	95	100	100	100.0	95005	PASS
96	95	5	9	6.7	6401	PASS
173	174	0.00	2	1.3	1151	PASS
174	95	50	100	90.7	86205	PASS
175	174	5	9	7.6	6521	PASS
176	174	95	101	95.1	81971	PASS
177	176	5	9	6.6	5395	PASS