

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041724\
 Data File : VY017951.D
 Acq On : 17 Apr 2024 09:45
 Operator : SY/MD
 Sample : VY0417SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0417SBL01

Quant Time: Apr 18 02:48:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	356285	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	638627	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	574648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	222294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	175997	57.083	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.160%
35) Dibromofluoromethane	7.716	113	203469	51.643	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.280%
50) Toluene-d8	10.179	98	742509	49.666	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.340%
62) 4-Bromofluorobenzene	12.483	95	209592	45.267	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.540%
Target Compounds						
20) Methylene Chloride	4.716	84	40361	9.779	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041724\
Data File : VY017951.D
Acq On : 17 Apr 2024 09:45
Operator : SY/MD
Sample : VY0417SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0417SBL01

Quant Time: Apr 18 02:48:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

