

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041724\
 Data File : VY017956.D
 Acq On : 17 Apr 2024 12:33
 Operator : SY/MD
 Sample : P2123-02RE
 Misc : 2.86g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RPI-SS-02RE

Quant Time: Apr 18 02:50:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	329507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	558674	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	418984	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	114549	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	150895	52.918	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.840%	
35) Dibromofluoromethane	7.722	113	179304	52.022	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.040%	
50) Toluene-d8	10.185	98	626601	47.911	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 95.820%	
62) 4-Bromofluorobenzene	12.483	95	126417	31.211	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 62.420%	
Target Compounds						Qvalue
20) Methylene Chloride	4.716	84	74635	19.553	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041724\
Data File : VY017956.D
Acq On : 17 Apr 2024 12:33
Operator : SY/MD
Sample : P2123-02RE
Misc : 2.86g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RPI-SS-02RE

Quant Time: Apr 18 02:50:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

