

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041824\
 Data File : VY017972.D
 Acq On : 18 Apr 2024 13:39
 Operator : SY/MD
 Sample : P2147-07
 Misc : 6.19g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 GRAB-3

Quant Time: Apr 19 02:00:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	103929	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	141843	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	100922	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.434	152	30403	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	30095	33.462	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	66.920%
35) Dibromofluoromethane	7.728	113	43325	49.509	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.020%
50) Toluene-d8	10.191	98	149648	45.068	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.140%
62) 4-Bromofluorobenzene	12.489	95	29529	28.714	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	57.420%
Target Compounds						
20) Methylene Chloride	4.716	84	8743	7.262	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041824\
Data File : VY017972.D
Acq On : 18 Apr 2024 13:39
Operator : SY/MD
Sample : P2147-07
Misc : 6.19g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
GRAB-3

Quant Time: Apr 19 02:00:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

