

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041921\
 Data File : VY004496.D
 Acq On : 19 Apr 2021 14:28
 Operator : SY/MD
 Sample : M2075-04
 Misc : 6.50G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1-D

Quant Time: Apr 20 01:48:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	140810	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	247227	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	237830	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	110715	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	94868	60.64	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 121.28%		
35) Dibromofluoromethane	7.728	113	76645	55.86	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 111.72%		
50) Toluene-d8	10.185	98	306305	52.49	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 104.98%		
62) 4-Bromofluorobenzene	12.483	95	107974	54.52	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 109.04%		
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	12814	8.48	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041921\
Data File : VY004496.D
Acq On : 19 Apr 2021 14:28
Operator : SY/MD
Sample : M2075-04
Misc : 6.50G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1-D

Quant Time: Apr 20 01:48:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

