

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041921\
 Data File : VY004501.D
 Acq On : 19 Apr 2021 16:25
 Operator : SY/MD
 Sample : M2075-19
 Misc : 6.26G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-6-S

Quant Time: Apr 20 01:49:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	138241	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	240954	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	215655	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	90525	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	84167	54.80	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.60%	
35) Dibromofluoromethane	7.728	113	71044	53.13	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.26%	
50) Toluene-d8	10.185	98	292335	51.40	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.80%	
62) 4-Bromofluorobenzene	12.483	95	93532	48.46	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041921\
Data File : VY004501.D
Acq On : 19 Apr 2021 16:25
Operator : SY/MD
Sample : M2075-19
Misc : 6.26G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-6-S

Quant Time: Apr 20 01:49:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

