

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041922\
 Data File : VY008394.D
 Acq On : 19 Apr 2022 14:33
 Operator : KP/MD
 Sample : N2427-01
 Misc : 6.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-1

Quant Time: Apr 20 06:28:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	85190	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	152479	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	134143	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	58602	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	47364	59.189	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.380%
35) Dibromofluoromethane	7.716	113	48523	55.096	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.200%
50) Toluene-d8	10.179	98	180431	59.873	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	119.740%
62) 4-Bromofluorobenzene	12.483	95	55837	49.262	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.520%
Target Compounds						
						Qvalue
16) Acetone	3.948	43	1721	8.942	ug/l	91
84) 1,2,4-Trimethylbenzene	13.123	105	7707	2.180	ug/l	98
95) Naphthalene	15.239	128	3845	1.613	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041922\
Data File : VY008394.D
Acq On : 19 Apr 2022 14:33
Operator : KP/MD
Sample : N2427-01
Misc : 6.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-1

Quant Time: Apr 20 06:28:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

