

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041922\
 Data File : VY008399.D
 Acq On : 19 Apr 2022 16:30
 Operator : KP/MD
 Sample : N2427-08
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-8

Quant Time: Apr 20 06:29:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 08 16:02:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	94881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	175031	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	172385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	80191	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	57369	64.370	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.740%
35) Dibromofluoromethane	7.716	113	56717	56.103	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.200%
50) Toluene-d8	10.179	98	214331	61.959	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	123.920%
62) 4-Bromofluorobenzene	12.483	95	73394	56.409	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY041922\
Data File : VY008399.D
Acq On : 19 Apr 2022 16:30
Operator : KP/MD
Sample : N2427-08
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-8

Quant Time: Apr 20 06:29:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y040822S.M
Quant Title : SW846 8260
QLast Update : Fri Apr 08 16:02:50 2022
Response via : Initial Calibration

