

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
 Data File : VY004538.D
 Acq On : 20 Apr 2021 18:51
 Operator : SY/MD
 Sample : M2098-01
 Misc : 5.61G/5ML/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Apr 21 08:54:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.795	168	121603	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	209432	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	194323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	92268	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	78241	57.91	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 115.82%			
35) Dibromofluoromethane	7.728	113	64154	55.20	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 110.40%			
50) Toluene-d8	10.185	98	258696	52.33	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 104.66%			
62) 4-Bromofluorobenzene	12.483	95	90372	53.87	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 107.74%			
Target Compounds					Qvalue	
11) Tert butyl alcohol	4.972	59	770	5.78	ug/l #	89
16) Acetone	3.960	43	2813	Below Cal		94
20) Methylene Chloride	4.735	84	6579	5.04	ug/l	96
31) Cyclohexane	7.801	56	3162	1.28	ug/l #	37
36) 1,1-Dichloropropene	7.795	75	10433	5.01	ug/l #	49
38) Carbon Tetrachloride	7.795	117	12193	6.78	ug/l #	17
52) Toluene	10.246	92	10891	2.88	ug/l	94
55) 1,1,2-Trichloroethane	10.618	97	2567	2.27	ug/l #	21
59) 2-Hexanone	10.801	43	4843	5.54	ug/l #	28
67) Ethyl Benzene	11.593	91	47582	6.41	ug/l	98
68) m/p-Xylenes	11.703	106	73244	26.57	ug/l	98
69) o-Xylene	12.032	106	54463	21.08	ug/l	98
73) Isopropylbenzene	12.331	105	24625	3.45	ug/l	100
74) N-amyl acetate	12.124	43	29313	12.21	ug/l #	63
75) 1,1,2,2-Tetrachloroethane	12.526	83	3848	2.85	ug/l #	26
76) 1,2,3-Trichloropropane	12.733	75	2216	2.24	ug/l #	100
78) n-propylbenzene	12.672	91	87492	10.39	ug/l	99
79) 2-Chlorotoluene	12.733	91	31951	6.94	ug/l #	41
80) 1,3,5-Trimethylbenzene	12.813	105	179876	31.04	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.483	75	54543	98.53	ug/l #	14
82) 4-Chlorotoluene	12.813	91	20696	4.25	ug/l #	51
83) tert-Butylbenzene	13.123	119	64034	12.91	ug/l #	64
84) 1,2,4-Trimethylbenzene	13.123	105	530557	90.72	ug/l	99
85) sec-Butylbenzene	13.251	105	54008	7.81	ug/l	83
86) p-Isopropyltoluene	13.373	119	46710	7.42	ug/l	99
89) n-Butylbenzene	13.697	91	79489	13.23	ug/l #	76
90) Hexachloroethane	13.977	117	23632	27.99	ug/l #	13
92) 1,2-Dibromo-3-Chloropr...	14.379	75	1013	4.18	ug/l #	31
95) Naphthalene	15.239	128	50499	13.26	ug/l #	75
96) 1,2,3-Trichlorobenzene	15.446	180	2145	1.39	ug/l #	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
Data File : VY004538.D
Acq On : 20 Apr 2021 18:51
Operator : SY/MD
Sample : M2098-01
Misc : 5.61G/5ML/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :

Quant Time: Apr 21 08:54:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.795 min Scan# 980
Delta R.T. -0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:168 Resp: 121603
Ion Ratio Lower Upper
168 100
99 57.4 45.5 68.3



#11
Tert butyl alcohol
Concen: 5.78 ug/l
RT: 4.972 min Scan# 517
Delta R.T. -0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion: 59 Resp: 770
Ion Ratio Lower Upper
59 100
57 5.8 8.1 12.1#









#16
Acetone
Concen: Below Cal
RT: 3.960 min Scan# 351
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 43 Resp: 2813
Ion Ratio Lower Upper
43 100
58 27.8 24.6 37.0



#20
Methylene Chloride
Concen: 5.04 ug/l
RT: 4.735 min Scan# 478
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51


Tgt Ion: 84 Resp: 6579
Ion Ratio Lower Upper
84 100
49 134.8 111.2 166.8
51 42.6 33.4 50.2
86 60.0 52.7 79.1








#31
Cyclohexane
Concen: 1.28 ug/l
RT: 7.801 min Scan# 981
Delta R.T. 0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51


Instrument :
MSVOA_Y
ClientSampled :





Tgt Ion: 56 Resp: 3162
Ion Ratio Lower Upper
56 100
69 112.4 25.8 38.6#
84 114.7 68.2 102.4#




#33
1,2-Dichloroethane-d4
Concen: 57.91 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. -0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51



Tgt Ion: 65 Resp: 78241
Ion Ratio Lower Upper
65 100
67 52.2 0.0 106.4









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:114 Resp: 209432
Ion Ratio Lower Upper
114 100
63 20.5 0.0 42.4
88 15.0 0.0 30.8



#35
Dibromofluoromethane
Concen: 55.20 ug/l
RT: 7.728 min Scan# 969
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion:113 Resp: 64154
Ion Ratio Lower Upper
113 100
111 102.7 82.6 123.8
192 19.3 16.1 24.1





#36

1,1-Dichloropropene

Concen: 5.01 ug/l

RT: 7.795 min Scan# 980





Delta R.T. -0.134 min

Lab File: VY004538.D

Acq: 20 Apr 2021 18:51

Tgt Ion: 75	Resp: 10433
Ion Ratio	Lower Upper
75 100	
110 8.0	17.3 51.7#
77 0.0	24.3 36.5#

Instrument :
MSVOA_Y
ClientSampled :



#38

Carbon Tetrachloride

Concen: 6.78 ug/l




RT: 7.795 min Scan# 980


Delta R.T. -0.115 min

Lab File: VY004538.D

Acq: 20 Apr 2021 18:51

Tgt Ion: 117	Resp: 12193
Ion Ratio	Lower Upper
117 100	
119 5.4	75.8 113.8#
121 0.0	25.3 37.9#









#50
Toluene-d8
Concen: 52.33 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 98 Resp: 258696
Ion Ratio Lower Upper
98 100
100 64.9 51.8 77.6



#52
Toluene
Concen: 2.88 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51


Tgt Ion: 92 Resp: 10891
Ion Ratio Lower Upper
92 100
91 179.8 137.8 206.6








#55
1,1,2-Trichloroethane
Concen: 2.27 ug/l
RT: 10.618 min Scan# 1443
Delta R.T. -0.030 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Instrument :
MSVOA_Y
ClientSampled :







Tgt Ion	Ratio	Lower	Upper
97	100		
83	12.5	68.4	102.6#
85	0.0	46.2	69.4#
99	0.0	48.8	73.2#



#59
2-Hexanone
Concen: 5.54 ug/l
RT: 10.801 min Scan# 1473
Delta R.T. -0.036 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion	Ratio	Lower	Upper
43	100		
58	1.0	25.4	76.4#









#62
4-Bromofluorobenzene
Concen: 53.87 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 95 Resp: 90372
Ion Ratio Lower Upper
95 100
174 82.4 0.0 162.4
176 80.4 0.0 156.0



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.490 min Scan# 1586
Delta R.T. -0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion: 117 Resp: 194323
Ion Ratio Lower Upper
117 100
82 56.7 44.9 67.3
119 31.5 25.4 38.0









#67
Ethyl Benzene
Concen: 6.41 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. -0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 91 Resp: 47582
Ion Ratio Lower Upper
91 100
106 30.3 25.1 37.7



#68
m/p-Xylenes
Concen: 26.57 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion:106 Resp: 73244
Ion Ratio Lower Upper
106 100
91 201.6 163.2 244.8









#69
o-Xylene
Concen: 21.08 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:106 Resp: 54463
Ion Ratio Lower Upper
106 100
91 218.1 107.4 322.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion:152 Resp: 92268
Ion Ratio Lower Upper
152 100
115 71.5 29.6 88.8
150 156.1 0.0 348.4





#73

Isopropylbenzene

Concen: 3.45 ug/l

RT: 12.331 min Scan# 1724

Delta R.T. 0.000 min

Lab File: VY004538.D





Acq: 20 Apr 2021 18:51

Tgt Ion:105 Resp: 24625

Ion Ratio Lower Upper

105 100

120 26.3 13.2 39.5



#74

N-ethyl acetate

Concen: 12.21 ug/l

RT: 12.124 min Scan# 1690

Delta R.T. -0.024 min

Lab File: VY004538.D

Acq: 20 Apr 2021 18:51

Tgt Ion: 43 Resp: 29313




Ion Ratio Lower Upper


43 100

70 9.4 30.4 45.6#

55 18.2 17.4 26.0

61 0.0 18.2 27.2#









#75
 1,1,2,2-Tetrachloroethane
 Concen: 2.85 ug/l
 RT: 12.526 min Scan# 1756
 Delta R.T. -0.061 min
 Lab File: VY004538.D
 Acq: 20 Apr 2021 18:51




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion: 83 Resp: 3848
 Ion Ratio Lower Upper
 83 100
 131 0.0 5.1 15.4#
 85 0.0 31.6 94.8#



#76
 1,2,3-Trichloropropane
 Concen: 2.24 ug/l
 RT: 12.733 min Scan# 1790
 Delta R.T. 0.098 min
 Lab File: VY004538.D
 Acq: 20 Apr 2021 18:51

Tgt Ion: 75 Resp: 2216
 Ion Ratio Lower Upper
 75 100
 77 1461.7 0.0 0.0#









#78
n-propylbenzene
Concen: 10.39 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 91 Resp: 87492
Ion Ratio Lower Upper
91 100
120 22.4 11.5 34.4



#79
2-Chlorotoluene
Concen: 6.94 ug/l
RT: 12.733 min Scan# 1790
Delta R.T. -0.024 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion: 91 Resp: 31951
Ion Ratio Lower Upper
91 100
126 0.0 16.7 50.1#









#80
 1,3,5-Trimethylbenzene
 Concen: 31.04 ug/l
 RT: 12.813 min Scan# 1803
 Delta R.T. -0.006 min
 Lab File: VY004538.D
 Acq: 20 Apr 2021 18:51




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion:105 Resp: 179876
 Ion Ratio Lower Upper
 105 100
 120 48.1 24.6 74.0



#81
 trans-1,4-Dichloro-2-butene
 Concen: 98.53 ug/l
 RT: 12.483 min Scan# 1749
 Delta R.T. 0.098 min
 Lab File: VY004538.D
 Acq: 20 Apr 2021 18:51

Tgt Ion: 75 Resp: 54543
 Ion Ratio Lower Upper
 75 100
 53 1.2 75.0 112.6#
 89 3.0 36.1 54.1#









#82
4-Chlorotoluene
Concen: 4.25 ug/l
RT: 12.813 min Scan# 1803
Delta R.T. -0.042 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion: 91 Resp: 20696
Ion Ratio Lower Upper
91 100
126 5.8 16.9 50.7#



#83
tert-Butylbenzene
Concen: 12.91 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.043 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion: 119 Resp: 64034
Ion Ratio Lower Upper
119 100
91 88.9 32.6 98.0
134 0.0 12.9 38.6#









#84
1,2,4-Trimethylbenzene
Concen: 90.72 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:105 Resp: 530557
Ion Ratio Lower Upper
105 100
120 45.4 22.4 67.3



#85
sec-Butylbenzene
Concen: 7.81 ug/l
RT: 13.251 min Scan# 1875
Delta R.T. -0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion:105 Resp: 54008
Ion Ratio Lower Upper
105 100
134 28.1 10.1 30.3









#86
p-Isopropyltoluene
Concen: 7.42 ug/l
RT: 13.373 min Scan# 1895
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:119 Resp: 46710
Ion Ratio Lower Upper
119 100
134 25.9 13.3 39.8
91 24.1 11.9 35.7



#89
n-Butylbenzene
Concen: 13.23 ug/l
RT: 13.697 min Scan# 1948
Delta R.T. -0.006 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion: 91 Resp: 79489
Ion Ratio Lower Upper
91 100
92 61.8 26.6 79.8
134 0.0 12.7 38.1#









#90
Hexachloroethane
Concen: 27.99 ug/l
RT: 13.977 min Scan# 1994
Delta R.T. 0.013 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51




Instrument :
MSVOA_Y
ClientSampled :


Tgt Ion:117 Resp: 23632
Ion Ratio Lower Upper
117 100
201 0.0 36.6 109.9#



#92
1,2-Dibromo-3-Chloropropane
Concen: 4.18 ug/l
RT: 14.379 min Scan# 2060
Delta R.T. 0.019 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion: 75 Resp: 1013
Ion Ratio Lower Upper
75 100
155 103.6 40.6 121.8
157 0.0 53.0 159.0#









#95
Naphthalene
Concen: 13.26 ug/l
RT: 15.239 min Scan# 2201
Delta R.T. 0.000 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Instrument :
MSVOA_Y
ClientSampled :

Tgt Ion:128 Resp: 50499
Ion Ratio Lower Upper
128 100
127 0.0 10.2 15.2#
129 17.2 8.8 13.2#



#96
1,2,3-Trichlorobenzene
Concen: 1.39 ug/l
RT: 15.446 min Scan# 2235
Delta R.T. 0.019 min
Lab File: VY004538.D
Acq: 20 Apr 2021 18:51

Tgt Ion:180 Resp: 2145
Ion Ratio Lower Upper
180 100
182 0.0 47.6 142.8#
145 1371.0 16.3 48.8#

